

RECEIVED

By Brian Lutmer at 9:17 am, Jan 16, 2015



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
DATAMASTER MAINTENANCE REPORT

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 201253	NAME OF AGENCY Miner Police Department	DATE OF INSPECTION 01/13/2015
-------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 103 State Highway H	TIME OF INSPECTION 4:07 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01-13-2015 16:07
--	--

<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
--	--

<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
---	---

<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 48 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
--	---

<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
---	---

<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
---	---

<input checked="" type="checkbox"/> INDICATOR LIGHTS
--

<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Repro</u>	LOT # <u>13002</u>	EXP. DATE <u>06/19/2015</u>
--	--------------------	-----------------------------

<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C	SIMULATOR SN <u>SD2273</u>	EXP. DATE <u>01/13/2016</u>
--	----------------------------	-----------------------------

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1  .102	TEST 2  .103	TEST 3  .103
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	1	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS  
(USE OTHER SIDE IF NECESSARY).

OPERATING WITHIN DOHSS STANDARDS

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME LANCE ASH
---------------	------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 240149 / 04/22/2016	TELEPHONE NUMBER (573) 471-8568
--	------------------------------------

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253  
01/13/15

TESTING OFFICER:  
ASH/LANCE  
OFFICER I.D.: 526  
PERMIT NUMBER: 240149  
EXPIRATION DATE: 04/22/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	16:13
INTERNAL STANDARD	VERIFIED	16:13
EXTERNAL STANDARD	.102	16:13
BLANK TEST	.000	16:14
EXTERNAL STANDARD	.103	16:13
BLANK TEST	.000	16:15
EXTERNAL STANDARD	.103	16:16
BLANK TEST	.000	16:17

N = 3  
SIN. = .1  
AVG. = .1026

Operator Signature

*Ash #526*

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253  
01/13/15  
16:07

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF0  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~"

Operator Signature

*Ash #526*

**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
MUIR POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253  
01/13/15

ARREST TIME: 16:02  
SUBJECT NAME:  
DOE/JOHN  
DOB: 12/22/45      SEX: M  
STATE/D.L.: MO/123456789  
ARRESTING OFFICER:  
ASH  
OFFICER I.D.: 526  
TESTING OFFICER:  
ASH/LANCE  
OFFICER I.D.: 526  
PERMIT NUMBER: 240149  
EXPIRATION DATE: 04/22/16  
MISCELLANEOUS DATA:  
N

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature

*Ash #526*

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER:** RepCo Marketing, Inc.  
**LOT NUMBER:** 13002  
**EXPIRATION DATE:** June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 13002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1217 gms/dl  $\pm$  .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.


This solution will produce a vapor alcohol value of .100  $\pm$  3% gms/210L Breath when heated to 34 Degrees Celsius  $\pm$  0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 20, 2013  
The expiration date for this lot number is June 19, 2015 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**

**JAMES L ASH**

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOXILYZER 5000**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2014

NUMBER 240149

EXPIRES 4/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-9731 (6-10)

LSAB-4106-101



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator ASH, JAMES  
Permit No 240149  
Date issued 4/22/2014 Date Expires 4/22/2016