MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:57 am, Aug 25, 2015

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance chec
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201253	MINER POLICE	08/21/2015
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
103 ST. HWY H, MINER, MO. 63801		8:29 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☑ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 08/21/15 08:29

☑ COMPUTER	☑ DETECTOR
☑ PROGRAM	☑ FILTERS
☑ HEATERS SAMPLE CHAMBER 49 °C	☑ QUARTZ STANDARD
☑ FLOW DETECTOR	☑ CALIBRATION
☑ PUMP HIGH SPEED	☑ PRINTER

☑ INDICATOR LIGHTS

☑ SIMULATOR SOLUTION SUPPLIER REPCO MARKETING LOT # 1400 EXP. DATE 04/30/2016

☑ SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD2273 EXP. DATE 01/13/2013

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☛ .100%	TEST 2 ☛ .100%	TEST 3 ☛ .102%

☑ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	2	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	2

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

OPERATING WITHIN DOHSS STANDARDS

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
▸	LANCE ASH
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240149 04/22/2016	(573) 471-8568

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **14001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1216** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **May 1, 2014**
The expiration date for this lot number is **April 30, 2016** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
08/21/15
08:29

---- DIAGNOSTIC CHECK ---

OMPUTER:	OKAY
ROGRAM (04-07-2009):	OKAY
ERTERS	
AMPLE CHAMBER:	49c
LOW DETECTOR:	OKAY
AMP	
IGH SPEED:	OKAY
ETECTOR:	OKAY
ILTER:	OKAY
UARTZ STANDARD:	OKAY
ALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Signature

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
08/21/15

TESTING OFFICER:
ASH
OFFICER I.D.: 521
PERMIT NUMBER: 248149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
M
N

---- SUPERVISOR MODE ---

BLANK TEST	.000	08:40
INTERNAL STANDARD	VERIFIED	08:40
EXTERNAL STANDARD	.100	08:40
BLANK TEST	.000	08:41
EXTERNAL STANDARD	.100	08:42
BLANK TEST	.000	08:42
EXTERNAL STANDARD	.100	08:43
BLANK TEST	.000	08:44

N = 3
STD. = .1
AVG. = .1006

Operator Signature

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MINER POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201253
08/21/15

ARREST TIME: 08:20
SUBJECT NAME:
DOE/JOHN
DOB: 12/24/55 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
ASH/L
OFFICER I.D.: 521
TESTING OFFICER:
ASH/L
OFFICER I.D.: 521
PERMIT NUMBER: 240149
EXPIRATION DATE: 04/22/16
MISCELLANEOUS DATA:
N
N

--- BREATH ANALYSIS ---

BLANK TEST	.000	08:48
INTERNAL STANDARD	VERIFIED	08:49
RADIO INTERFERENCE		

Operator Signature ASH

2208-02

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JAMES L ASH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOXILYZER 5000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2014

NUMBER 240149

EXPIRES 4/22/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

RO 580-077 (6-10) LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ASH, JAMES
Permit No 240149
Date Issued 4/22/2014 Date Expires 4/22/2016