



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Brian Lutmer at 4:44 pm, Jan 20, 2015

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

| | | |
|---|--|---|
| DATAMASTER SN <u>201286</u> | NAME OF AGENCY <u>Mtn. Grove Police</u> | DATE OF INSPECTION <u>01-20-2015</u> |
| LOCATION OF INSTRUMENT (STREET AND CITY) <u>102 E. STATE ST. Mtn Grove</u> | | TIME OF INSPECTION <u>0838</u> |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

| | |
|--|---|
| <input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED) | DATE AND TIME (from printout) <u>01-20-15 0839</u> |
| <input checked="" type="checkbox"/> COMPUTER | <input checked="" type="checkbox"/> DETECTOR |
| <input checked="" type="checkbox"/> PROGRAM | <input checked="" type="checkbox"/> FILTERS |
| <input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49</u> °C | <input checked="" type="checkbox"/> QUARTZ STANDARD |
| <input checked="" type="checkbox"/> FLOW DETECTOR | <input checked="" type="checkbox"/> CALIBRATION |
| <input checked="" type="checkbox"/> PUMP HIGH SPEED | <input checked="" type="checkbox"/> PRINTER |

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER GLWTH LOT # 140030 EXP. DATE 01/20/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C SIMULATOR SN SD3322 EXP. DATE 12/16/15

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 * .099 TEST 2 * .098 TEST 3 * .098

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | |
|----------|----------|-----------|-----------|-----------|----------|
| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) | (.15-.19) | OVER .19 |
| <u>1</u> | <u>1</u> | <u>0</u> | <u>0</u> | <u>0</u> | <u>1</u> |

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

| | |
|---|---|
| SIGNATURE <u>Daniel R. Bledsoe</u> | PRINT FULL NAME <u>Daniel R. Bledsoe</u> |
| TYPE II PERMIT NUMBER/EXPIRATION DATE <u>840433 12/17/2016</u> | TELEPHONE NUMBER <u>417.926.5181</u> |

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/20/15

ARREST TIME: 08:30
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/000000000
ARRESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.: 1201
TESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.: 1201
PERMIT NUMBER: 240433
EXPIRATION DATE: 12/17/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

| | | | |
|--------------------------|-------------------|----------|-------|
| <input type="checkbox"/> | BLANK TEST | .000 | 08:49 |
| <input type="checkbox"/> | INTERNAL STANDARD | VERIFIED | 08:49 |
| <input type="checkbox"/> | SUBJECT SAMPLE | .000 | 08:49 |
| <input type="checkbox"/> | BLANK TEST | .000 | 08:50 |

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT


BAC DATAMASTER SERIAL NUMBER 201286
01/20/15

ARREST TIME: 08:30
SUBJECT NAME:
RFI/TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/000000000
ARRESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.: 1201
TESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.: 1201
PERMIT NUMBER: 240433
EXPIRATION DATE: 12/17/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

| | | | |
|--------------------------|--------------------|----------|-------|
| <input type="checkbox"/> | BLANK TEST | .000 | 08:47 |
| <input type="checkbox"/> | INTERNAL STANDARD | VERIFIED | 08:47 |
| <input type="checkbox"/> | RADIO INTERFERENCE | | |

Operator Signature



220

Operator Signature



2208-02

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/20/15

TESTING OFFICER:
BLEDSOE/DANNY
OFFICER I.D.# 1201
PERMIT NUMBER: 240433
EXPIRATION DATE: 12/17/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

| | | |
|-------------------|----------|-------|
| BLANK TEST | .000 | 08:41 |
| INTERNAL STANDARD | VERIFIED | 08:41 |
| EXTERNAL STANDARD | .099 | 08:42 |
| BLANK TEST | .000 | 08:42 |
| EXTERNAL STANDARD | .099 | 08:43 |
| BLANK TEST | .000 | 08:43 |
| EXTERNAL STANDARD | .099 | 08:44 |
| BLANK TEST | .000 | 08:44 |

n = 3
STDEV = .1
HYG. = .0996

Face This Side Down - This Edge In First

**BAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
MOUNTAIN GROVE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
01/20/15
08139

--- DIAGNOSTIC CHECK ---

| | |
|-----------------------|------|
| COMPUTER: | OKAY |
| PROGRAM (04-87-2009): | OKAY |
| HEATERS | |
| SAMPLE CHAMBER: | 49c |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFGB
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ?

Operator Signature

2208-02

Operator Signature

2206



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14030 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on January 22, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1215% (w/vol) ethyl alcohol. The expiration date for this lot number is January 20, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DANIEL BLEDSOE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/17/2014

NUMBER 240433

EXPIRES 12/17/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dan Voshell
acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (16-10)

MO 560-0771 (16-10)