MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
By Brian Lutmer at 3:02 pm, Nov 03, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
201288	Campbell Police Dept.	11/01/2015

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
204 W. Grand Campbell MO 63933	~~7:02 am~~ 7:32am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 11/01/15 07:32
  - [x] COMPUTER
  - [x] DETECTOR
  - [x] PROGRAM
  - [x] FILTERS
  - [x] HEATERS SAMPLE CHAMBER 49°C
  - [x] QUARTZ STANDARD
  - [x] FLOW DETECTOR
  - [x] CALIBRATION
  - [x] PUMP HIGH SPEED
  - [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER RepCo LOT # 15001 EXP. DATE 05/20/2017
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD174301 EXP. DATE 01/14/2016
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] ~~0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE~~
- [ ] ~~0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE~~

TEST 1	TEST 2	TEST 3
.097%	>097%	.098%

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Instrument is within D.O.H. specifications. Simulator soution RepCo lot # 15001 exp. date 05-20-17. Vapor alcohol value .100%+/- 3%

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Weldon Wallace	Weldon Wallace
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
240323 08/12/2016	(573) 246-2511

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 15001
EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **15001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1206** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **May 21, 2015** The expiration date for this lot number is **May 20, 2017** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

WELDON WALLACE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/12/2014

NUMBER 240323

EXPIRES 8/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WALLACE, WELDON
Permit No 240323
Date Issued 8/12/2014 **Date Expires** 8/12/2016

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201288
11/01/15

TESTING OFFICER:
WALLACE/WELDON/L
OFFICER I.D.: 23
PERMIT NUMBER: 240323
EXPIRATION DATE: 03/12/16
MISCELLANEOUS DATA:
MAINTENANCE CHECK

---- SUPERVISOR MODE ----

BLANK TEST	.000	07:36
INTERNAL STANDARD	VERIFIED	07:36
EXTERNAL STANDARD	.097	07:36
BLANK TEST	.000	07:37
EXTERNAL STANDARD	.097	07:37
BLANK TEST	.000	07:38
EXTERNAL STANDARD	.098	07:38
BLANK TEST	.000	07:39

N = 3
STA. = .1
AVG. = .0973

RATOR SIGNATURE

Stock No.
1

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201288
11/01/15
07:36

---- DIAGNOSTIC CHECK ----

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
CAMPBELL POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 201286
11/01/15

ARREST TIME: 07:00
SUBJECT NAME:
DOE/JOHN
DOB: 10/10/70 SEX: M
STATE/D.L.: MO/123456
ARRESTING OFFICER:
WALLACE/WELDON/L
OFFICER I.D.: 23
TESTING OFFICER:
WALLACE/WELDON/L
OFFICER I.D.: 23
PERMIT NUMBER: 240623
EXPIRATION DATE: 08/12/16
MISCELLANEOUS DATA:
MAINTENANCE CHECK

--- BREATH ANALYSIS ---

BLANK TEST	.000	07:44
INTERNAL STANDARD	VERIFIED	07:44
RADIO INTERFERENCE		

OPERATOR SIGNATURE

Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901