



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
 By Carol Day at 10:07 am, Jun 25, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days since last check).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>203005</u>	NAME OF AGENCY <u>Edmundson P.D.</u>	DATE OF INSPECTION <u>6/15/15</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>4430 Holman Ln, Edmundson, MO 63134</u>		TIME OF INSPECTION <u>1110</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>6/15/15 1113</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>50</u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>both</u>	LOT # <u>14720</u> EXP. DATE <u>9/24/16</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34.0</u> °C	SIMULATOR SN <u>SD2737</u> EXP. DATE <u>7/15/15</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>.098%</u>	TEST 2 <u>.097%</u>	TEST 3 <u>.097%</u>
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u> (0-.04)	<u>0</u> (.05-.09)	<u>0</u> (.10-.14)	<u>0</u> (.15-.19)	<u>0</u> OVER .19
---------------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE <u>[Signature]</u>	PRINT FULL NAME <u>Sgt - Daniel K. Patrick #20</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>240040 2/24/16</u>	TELEPHONE NUMBER <u>314-428-4577</u>

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 203085
06/15/15
11:13


--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HEATERS
SAMPLE CHAMBER: 50c
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnop
qrstuvwxyz{|}~

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 203005
06/15/15

ARREST TIME: 11:00
SUBJECT NAME:
DOE/JOHN
DOB: 01/01/82 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
PATRICK/DANIEL
OFFICER I.D.: 20
TESTING OFFICER:
PATRICK/DANIEL
OFFICER I.D.: 20
PERMIT NUMBER: 240040
EXPIRATION DATE: 02/24/16
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE

--- BREATH ANALYSIS ---

BLANK TEST	.000	11:18
INTERNAL STANDARD	VERIFIED	11:18
RADIO INTERFERENCE		

DR SIGNATURE



* No.

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 203005
06/15/15

TESTING OFFICER:
PATRICK/DANIEL
OFFICER I.D.: 20
PERMIT NUMBER: 240040
EXPIRATION DATE: 02/24/16
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE

--- SUPERVISOR NODE ---

BLANK TEST	.000	11:28
INTERNAL STANDARD	VERIFIED	11:28
EXTERNAL STANDARD	.095	11:29
BLANK TEST	.000	11:29
EXTERNAL STANDARD	.097	11:30
BLANK TEST	.000	11:30
EXTERNAL STANDARD	.097	11:31
BLANK TEST	.000	11:31

N = 3
SIM. = .1
AVG. = .0973

OPERATOR SIGNATURE



Card Stock No.
60021

REORDER ALL SUPPLIES FROM N.P.A.S.
P.O. BOX 1435, MANSFIELD, OH 44901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DANEIL R PATRICK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/24/2014

NUMBER 240040

EXPIRES 2/24/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES