MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 11:38 am, Mar 24, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
203048	MALDEN POLICE DEPARTMENT	03-21-15
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
MALDEN POLICE DEPARTMENT 112 E. LACLEDE MALDEN MO 63863		1200

CHECK LIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 03-21-15 1219
- ☒ COMPUTER — ☒ DETECTOR
- ☒ PROGRAM — ☒ FILTERS
- ☒ HEATERS SAMPLE CHAMBER 50 °C — ☒ QUARTZ STANDARD
- ☒ FLOW DETECTOR — ☒ CALIBRATION
- ☒ PUMP HIGH SPEED — ☒ PRINTER
- ☒ INDICATOR LIGHTS
- ☒ SIMULATOR SOLUTION SUPPLIER GUTH LABORATORIES LOT # 13290 EXP. DATE 10-29-15
- ☒ SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD2589 EXP. DATE 06-02-15
- ☒ CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.102	.102	.100

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

Calibration Performed on Instrument

Working Within Department of Health Guidelines

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Russell Miller	Sgt. Russell Miller
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240233 05-12-16	573-276-2211

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MALDEN POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 203048
03/21/15

ARREST TIME: 00:01
SUBJECT NAME:
BLANK/TEST
DOB: 01/01/11 SEX: M
STATE/D.L.: MO/1234567
ARRESTING OFFICER:
MILLER/RUSSELL/L
OFFICER I.D.: 3
TESTING OFFICER:
MILLER/RUSSELL/L
OFFICER I.D.: 3
PERMIT NUMBER: 240233
EXPIRATION DATE: 05/12/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	13:03
INTERNAL STANDARD	VERIFIED	13:03
SUBJECT SAMPLE	.000	13:03
BLANK TEST	.000	13:04

Operator Signature

2206-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MALDEN POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 203048
03/21/15

TESTING OFFICER:
MILLER/RUSSELL/L
OFFICER I.D.: 3
PERMIT NUMBER: 240233
EXPIRATION DATE: 05/12/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	12:46
INTERNAL STANDARD	VERIFIED	12:46
EXTERNAL STANDARD	.100	12:47
BLANK TEST	.000	12:47
EXTERNAL STANDARD	.100	12:48
BLANK TEST	.000	12:49
EXTERNAL STANDARD	.100	12:49
BLANK TEST	.000	12:50

[illegible] = .3
[illegible] = .1
[illegible] = .1913

Operator Signature

Mar 21 2015 01:16pm P004/006

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MALDEN POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 208048
03/21/15
12:19

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	50c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

```
 !"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~
```

Operator Signature

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
MALDEN POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 208048
03/21/15

ARREST TIME: 00:01
SUBJECT NAME:
TEST/ONE/
DOB: 01/01/11 SEX: M
STATE/D.L.: MO/1234567
ARRESTING OFFICER:
MILLER/RUSSELL/L
OFFICER I.D.: 3
TESTING OFFICER:
MILLER/RUSSELL/L
OFFICER I.D.: 3
PERMIT NUMBER: 248233
EXPIRATION DATE: 05/12/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	12:58
INTERNAL STANDARD	VERIFIED	12:58
RADIO INTERFERENCE		

Operator Signature

Mar 21 2015 01:16pm P005/006

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

RUSSELL L MILLER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/12/2014

NUMBER 240233

EXPIRES 5/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICE

MO 580-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MILLER, RUSSELL
Permit No 240233
Date Issued 5/12/2014 Date Expires 5/12/2016