MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

RECEIVED REPORT #6
By Carol Day at 8:11 am, Aug 04, 2015

omplete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 d
omplete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
etain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

ATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
:03052	MOUNTAIN VIEW POLICE	07/27/2015
OCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
17 NORTH PINE STREET MOUNTAIN VIEW MO 65548		4:21 pm

HECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values here determined.) Unmarked items must be corrected before using instrument.

] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) DATE AND TIME (from printout) 07/27/2015 17:00

- ☑ COMPUTER
- ☑ DETECTOR
- ☑ PROGRAM
- ☑ FILTERS
- ☑ HEATERS SAMPLE CHAMBER 49 °C
- ☑ QUARTZ STANDARD
- ☑ FLOW DETECTOR
- ☑ CALIBRATION
- ☑ PUMP HIGH SPEED
- ☑ PRINTER

] INDICATOR LIGHTS

] SIMULATOR SOLUTION SUPPLIER GUTH LOT # 14110 EXP. DATE 05/01/2016

] SIMULATOR TEMP (34°C ± 0.2°C) 34.02 °C SIMULATOR SN SD1425 EXP. DATE 06/30/2016

] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

EST 1 ☛ .096	TEST 2 ☛ .097	TEST 3 ☛ .097

] PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

IDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:)O NOT INCLUDE SELF-ADMINISTERED TESTS)

EFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	1	0	0

T ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS E OTHER SIDE IF NECESSARY).

IONTHLY MAINTANCE INSTRUTMENT REPLACES OTHER DATAMASTER AS IT HAS BEEN TAKEN OUT OF ERVICE. MONTHLY MAINTANCE ON CURRENT INSTRUTMENT DONE INSTRUTMENT FUNCTION WITH NO ROBLMENS. INSTRUTMENT OK

SPECTING OFFICER

NATURE	PRINT FULL NAME
Thomas R Poindexter	THOMAS R POINDEXTER
E II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
0213 04/30/2016	(417) 934-2525

URN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

GUTH LABORATORIES, INC.
[illegible] • [illegible] • [illegible]

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

ST Traceability:
sting was conducted using Cerilliant Reference Standard lot number PN122211-02 whose
lues are traceable to NIST.
balances are calibrated annually by an outside agency using NIST traceable weights.
libration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
THOMAS R POINDEXTER

reby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, operate the following breath analyzer(s):

DATAMASTER

ne determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

E 4/30/2014

MBER 240213

PIRES 4/30/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-10)

0-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator POINDEXTER, THOMAS
Permit No 240213
Date Issued 4/30/2014 Date Expires 4/30/2016

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 293052
07/27/15

TESTING OFFICER:
POINDEXTER/THOMAS/K
OFFICER I.D.: 122
PERMIT NUMBER: 240213
EXPIRATION DATE: 04/30/16
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	16:24
INTERNAL STANDARD	VERIFIED	16:24
EXTERNAL STANDARD	.096	16:25
BLANK TEST	.000	16:25
EXTERNAL STANDARD	.097	16:26
BLANK TEST	.000	16:26
EXTERNAL STANDARD	.097	16:27
BLANK TEST	.000	16:27

N = 3
STN. = .098
AVG. = .0966

Operator Signature Thomas R Poindexter

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
BAC DATAMASTER SERIAL NUMBER 293052
07/27/15

ARREST TIME: 16:00
SUBJECT NAME:
RADIO/TEST
DOB: 09/18/67 SEX: M
STATE/D.L.: A2/85222
ARRESTING OFFICER:
POINDEXTER/THOMAS/K
OFFICER I.D.: 122
TESTING OFFICER:
POINDEXTER/THOMAS/K
OFFICER I.D.: 122
PERMIT NUMBER: 240213
EXPIRATION DATE: 04/30/16
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	16:31
INTERNAL STANDARD	VERIFIED	16:32
RADIO INTERFERENCE		

Operator Signature Thomas R Poindexter

2208-02