



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 REPORT #6
 By Carol Day at 1:17 pm, Sep 09, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 3 months) or whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 203052	NAME OF AGENCY MOUNTAIN VIEW POLICE	DATE OF INSPECTION 09/01/2017
LOCATION OF INSTRUMENT (STREET AND CITY) 117 NORTH PINE STREET		TIME OF INSPECTION 10:15

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 09/01/2017 10:45
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER guth	LOT # 14110 EXP. DATE 05/01/2016
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.02 °C	SIMULATOR SN sd1425 EXP. DATE 06/30/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .097.	TEST 2 .097	TEST 3 .098
--------------	-------------	-------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	1	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).
 monthly maintance instrutment ok

INSPECTING OFFICER	
SIGNATURE <i>Thomas R. Poindexter</i>	PRINT FULL NAME thomas .r. poindexter
TYPE II PERMIT NUMBER/EXPIRATION DATE 250167 07/23/2017	TELEPHONE NUMBER (417) 934-2525

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.
ESTABLISHED 1957 • MEMPHIS, TENNESSEE • 901-944-1111 • TELEFAX 901-944-1112

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14110 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 5, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL SN: 610N9036269, and found to contain 0.1206% (w/vol) ethyl alcohol. The expiration date for this lot number is May 1, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number #R022211-02 whose mass are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

THOMAS R POINDEXTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/23/2015

NUMBER 250167

EXPIRES 7/23/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (P6-10)

MO 580-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator POINDEXTER, THOMAS
Permit No 250167
Date Issued 7/23/2015 Date Expires 7/23/2017

Face This Side Down - This Edge In First

RAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 240052

09/01/15

10:15

----- DIAGNOSTIC CHECK -----

COMPUTER: OKAY

PROGRAM (04-07-2009): OKAY

HEATERS

SAMPLE CHAMBER: 49C

FLOW DETECTION: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstu vwxyz{|}~"

Signature Thomas R Poincloux

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI

BAC DATAMASTER SERIAL NUMBER 240052

09/01/15

TESTING OFFICER:

POINDEXTER/THOMAS/R

OFFICER I.D.#: 122

PERMIT NUMBER: 250167

EXPIRATION DATE: 07/23/17

MISCELLANEOUS DATA:

----- SUPERVISOR MODE -----

BLANK TEST	.000	10:18
INTERNAL STANDARD	.000 (ED)	10:19
EXTERNAL STANDARD	.000	10:19
BLANK TEST	.000	10:19
EXTERNAL STANDARD	.000	10:20
BLANK TEST	.000	10:21
EXTERNAL STANDARD	.000	10:21
BLANK TEST	.000	10:22

N = 3

SIG. = .000

AVG. = .0073

Operator Signature Thomas R Poincloux

2208-02

Face This Side Down - This Edge In First

**RAC DataMaster
Evidence Ticket**

STATE OF MISSOURI
RAC DATAMASTER SERIAL NUMBER 280682
09/01/15

ARREST TIME: 11:00
SUBJECT NAME:
WROTH/LEST
DOB: 09/18/76 SEX: M
STATE/D.O.B.: MO/09/76
ARRESTING OFFICER:
POINDEXTER/THOMPSON
OFFICER I.D.: 182
TESTING OFFICER:
POINDEXTER/THOMPSON
OFFICER I.D.: 182
PERMIT NUMBER: 050167
EXPIRATION DATE: 07/23/17
MISCELLANEOUS DATA:

--- BREATH ANALYSTS ---

BLANK TEST	LAB	10/15
INTERNAL STANDARD	VERT: 100	10.00
RADIO INTERFERENCE		

Operator Signature

Thomas R Poinc