MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to ex
Complete this report whenever the instrument is serviced or repaired and whenever it is placed i
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

RECEIVED
By Carol Day at 11:32 am, Apr 10, 2015

INTOX DMT SN	NAME OF AGENCY	DATE OF INSPECTION
500043	Missouri State Highway Patrol	04/01/2015
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
2990 NW Vivion Road, Riverside, MO. 64150		21:40:38

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

☒ DIAGNOSTIC RECORD

DATE AND TIME 04/01/2015 21:40:40

☒ PROGRAM

☒ SAMPLE CHAMBER 48.9°C

☒ BREATH TUBE 45.2°C

☒ PUMP

☒ DETECTOR

☒ FILTER 1

☒ FILTER 2

☒ FILTER 3

☒ INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

☐ SIMULATOR STANDARD

☒ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER ILMO LOT # 17513080A1 EXP. DATE 07/01/2015

☐ SIMULATOR TEMP (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

☐ 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

☒ 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

☐ 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.077	TEST 2: 0.076	TEST 3: 0.076

☒ PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 0	0-.04: 0	.05-.09: 1	.10-.14: 0	.15-.19: 0	OVER .19: 0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER

SIGNATURE		PRINT FULL NAME
RJ Wilhoit		RYAN J WILHOIT
TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
240081	03/07/2016	816-622-0800

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services
Southeast District Office
2875 James Blvd, Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis