

RECEIVED

By Brian Lutmer at 4:40 pm, Jan 21, 2015



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program DHSS

INTOX DMT SN 500219	NAME OF AGENCY Ozark Police Dept	DATE OF INSPECTION 01/21/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 201 E. Brick, Ozark, MO		TIME OF INSPECTION 14:57:00

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC RECORD	
DATE AND TIME <u>01/21/2015 14:57:02</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.8°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>48.0°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR STANDARD	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u>	LOT # <u>AG41890</u> EXP. DATE <u>07/08/2016</u>
<input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C)	SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.100	TEST 2: 0.099	TEST 3: 0.098
---------------	---------------	---------------

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 0	0-.04: 0	.05-.09: 0	10-.14: 2	.15-.19: 0	OVER .19: 1
-------------	----------	------------	-----------	------------	-------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

January 21, 2015 Maintenance test.

INSPECTING OFFICER

SIGNATURE <i>Scott D Stopka</i>	PRINT FULL NAME SCOTT D STOPKA
---------------------------------	-----------------------------------

TYPE II PERMIT NUMBER 240379	EXPIRATION DATE 10/17/2016	TELEPHONE NUMBER 417-581-7914
---------------------------------	-------------------------------	----------------------------------

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, MO Department of Health and Senior Services
Southeast District Office
2875 James Blvd, Poplar Bluff, MO 63901



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-73 8

Certificate of Analysis

Customer Name
 Intoximeters, Inc.
 081 Craig Road
 St. Louis, Mo 63146

Test Date: 8-Jul- 014

Lot # AG41890

<u>Exp Date</u> 8-Jul- 016	<u>Cyl Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.100 ± % BrAC (60 ppm) Balance
--------------------------------------	-------------------------------	--	--

Certification Traceable to N I S T RGM Ethanol Standards:

<u>Serial No</u>	<u>Concentration</u>	<u>Serial No</u>	<u>Concentration</u>
EB0010581	391 8 ppm	EB0010603	392 5 ppm
EB0010570	259 8 ppm	EB0010559	258 9 ppm
EB0010285	209 0 ppm	EB0010595	208 9 ppm
EB0010561	103 7 ppm	EB0010562	104 9 ppm
EB0010681	52 22 ppm	EB0010579	52 94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2014.07.08 16:23:57 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Analyst: _____

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SCOTT STOPKA


is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/17/2014
NUMBER 240379
EXPIRES 10/17/2016



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


,acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator STOPKA, SCOTT
Permit No 240379
Date issued 10/17/2014 Date Expires 10/17/2016