



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
and whenever instrument is
By Carol Day at 10:47 am, Dec 14, 2015

Complete this report in duplicate at the time of the regular monthly preventive maintenance check and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in de

INSTRUMENT SERIAL NUMBER 80-005841	LOCATION OF INSTRUMENT KCMO POLICE DEPT.	DATE OF INSPECTION 12/08/2015	TIME OF INSPECTION 22:13
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	22:17	DRY	05514080A1	04/01/2016
Cal Check	0.079	22:17	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	22:17	N/A	N/A	N/A
Cal Check	0.079	22:18	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	22:18	0.080	CMI	
Cal Check	0.079	22:19	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	22:19	0.079		
Cal Check	0.079	22:19	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	22:19	0.079		
Pass			CALIBRATION CHECK RESULT 3		
			0.079		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Test	Pass/Fail		Test	g/210L	Time
Voltage/Current Test	Pass		Air Blank	0.000	22:20
RAM Test	Pass		Subject Test	RFI*	22:20
EEPROM Checksum Test	Pass		Air Blank	0.000	22:20
Real Time Clock Test	Pass		*RFI Detect		
DSP Test	Pass		Pass		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
INSTRUMENT TESTED AND CERTIFIED WITHIN MISSOURI DOHSS GUIDELINES

INSPECTING OFFICER		
SIGNATURE <i>Tim Fillpot</i>	PRINT NAME TIM FILLPOT	
TYPE II PERMIT NUMBER 250125	EXPIRATION DATE 06/18/2017	TELEPHONE NUMBER 8164828195



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TIMOTHY L FILLPOT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015
NUMBER 250125
EXPIRES 6/8/2017

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (08-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator FILLPOT, TIMOTHY
Permit No. 250125
Date Issued 6/8/2015 Date Expires 6/8/2017



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmo-products.com

Certificate of Analysis

Certificate ID: 6358
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 05514080A1
Expiration: 4/1/2016

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	298 ppm	+/- 0.002 or 2% BAC whichever is greater	NDIR
Nitrogen	Balance		

*NIST Standard Reference Material
Cylinder No. CCI4290 / Job No. 09160202
Certified 212.8 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

03/24/14
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com



ISO/IEC 17025:2005 Accredited Laboratory