MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
By Carol Day at 11:22 am, Aug 07, 2015

REPORT #2

Complete this report in duplicate at the time of the regular monthly preve[illegible] repaired. Send one copy to Department of Health and Senior Services, [illegible] and retain one copy in department file.

INSTRUMENT SERIAL NUMBER	LOCATION OF INSTRUMENT	DATE OF INSPECTION	TIME OF INSPECTION
80-005844	KCMO POLICE DEPT.	08/05/2015	03:03

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	03:05
Cal Check	0.077	03:06
Air Blank	0.000	03:06
Cal Check	0.077	03:07
Air Blank	0.000	03:07
Cal Check	0.077	03:07
Air Blank	0.000	03:08

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
DRY	05514080A1	04/01/2016
SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
N/A	N/A	N/A
STANDARD VALUE	STANDARD SUPPLIER	
0.080	CMI	

CALIBRATION CHECK RESULT 1: 0.077

CALIBRATION CHECK RESULT 2: 0.077

CALIBRATION CHECK RESULT 3: 0.077

MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)
3.7%	0.000

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	03:08
Subject Test	RFI*	03:09
Air Blank	0.000	03:09

*RFI Detect

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
3	3	3	4	3	3

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Kori Smeiska	KORI SMEISKA

TYPE II PERMIT NUMBER	EXPIRATION DATE	TELEPHONE NUMBER
250133	06/08/2017	8162345000

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KORI SMEISKA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015

NUMBER 250133

EXPIRES 6/8/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SMEISKA, KORI
Permit No 250133
Date Issued 6/8/2015 Date Expires 6/8/2017

Cyl 059

ILMO specialty gases

7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 6358
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 05514080A1
Expiration: 4/1/2016

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208 ppm	+/- 0.002 or 2% BAC whichever is greater	NDIR
Nitrogen	Balance		

*NIST Standard Reference Material
Cylinder No. CC14290 / Job No. 09160202
Certified 212.8 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

03/24/14
Date

ISO/IEC 17025:2005 Accredited Laboratory

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

ISO/IEC 17025:2005 Accredited Laboratory