



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 9:31 am, Apr 29, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 931097	NAME OF AGENCY Lawson Police Department	DATE OF INSPECTION 4-20-2015
LOCATION OF INSTRUMENT (STREET AND CITY) 104 W Third Street, Lawson MO 64062		TIME OF INSPECTION 0553

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 04/20/15 05:53
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Repro Marketing Inc** LOT # **14001** EXP. DATE **4-30-16**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **DR5392** EXP. DATE **01-14-2016**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.098	TEST 2 0.098	TEST 3 0.098
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS —	(0-.04) —	(.05-.09) —	(.10-.14) —	(.15-.19) —	OVER .19 —
------------	-----------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS
 (USE OTHER SIDE IF NECESSARY).

Conforms with DHSS standards

INSPECTING OFFICER	
SIGNATURE Sgt. Bruce Summa	PRINT FULL NAME Sgt. Bruce Summa
TYPE PERMIT NUMBER/EXPIRATION DATE 10-03-2016 240362	TELEPHONE NUMBER

RETURN COMPLETED REPORT TO THE:
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 14001

EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 931097
04/20/15
05:53

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (04-07-2009): OKAY
HERTERS
SAMPLE CHAMBER: 492
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST
!#####()**+-./0123456789:;<=>?@ABCDEFGHIJK
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
P~p~s~t~u~v~w~x~y~z~()~*~+~

Operator Signature

Sgt Bruce Summers

2209-4

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 931097
04/20/15

PRETEST TIME: 05:00

SUBJECT NAME:
RPI/TEST

SEX: M

DOB: 12/30/56

STATE/D.L.I.: MO/123456

ARRESTING OFFICER:
SUMMR/BRUCE/R

OFFICER I.D.: 201

TESTING OFFICER:
SUMMR/BRUCE/R

OFFICER I.D.: 201

SUMMR/BRUCE/R

OFFICER I.D.: 201

PERMIT NUMBER: 240362

EXPIRATION DATE: 10/02/16

MISCELLANEOUS DATA:
RPI TEST

--- BREATH ANALYSIS ---

BLANK TEST .000 06:04
INTERNAL STANDARD VERIFIED 06:04
RADIO INTERFERENCE

Operator Signature

Sgt Bruce Summers

2209-4

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
LAWSON POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 931097
04/20/15

TESTING OFFICER:
SUMMR/BRUCE/R

OFFICER I.D.: 201

PERMIT NUMBER: 240362

EXPIRATION DATE: 10/02/16

MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST .000 05:56
INTERNAL STANDARD VERIFIED 05:56
EXTERNAL STANDARD .098 05:56
BLANK TEST .000 05:57
EXTERNAL STANDARD .098 05:57
BLANK TEST .000 05:58
EXTERNAL STANDARD .098 05:58
BLANK TEST .000 05:59

N = 3
SIM. = .1
RWS. = .298

Operator Signature

Sgt Bruce Summers

2209-02



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRUCE SUMMA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014
 NUMBER 240362
 EXPIRES 10/3/2016

MO 580-0771 (6-10)


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
Darl Vosterby
 ,acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SUMMA, BRUCE
 Permit No 240362
 Date Issued 10/3/2014 Date Expires 10/3/2016