MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED
By Carol Day at 12:39 pm, Apr 23, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN	NAME OF AGENCY	DATE OF INSPECTION
940097	Sweet Springs Police Dept	4/15/15
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
324 S. Miller Sweet Springs		13:40

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIAGNOSTIC CHECK (PRINTOUT ATTACHED) — DATE AND TIME (from printout) 04/15/15 13:40
- [x] COMPUTER
- [x] DETECTOR
- [x] PROGRAM
- [x] FILTERS
- [x] HEATERS SAMPLE CHAMBER 49 °C
- [x] QUARTZ STANDARD
- [x] FLOW DETECTOR
- [x] CALIBRATION
- [x] PUMP HIGH SPEED
- [x] PRINTER
- [x] INDICATOR LIGHTS
- [x] SIMULATOR SOLUTION SUPPLIER Rep Co Marketing LOT # 13002 EXP. DATE 6/19/2015
- [x] SIMULATOR TEMP (34°C ± 0.2°C) 34°C ± 0.2° °C SIMULATOR SN SD3144 EXP. DATE 08/25/15
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ☞	TEST 2 ☞	TEST 3 ☞
.099%	.100%	.099%

- [x] PERFORM R.F.I. TEST (PRINTOUT ATTACHED) RFI detected

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
0	0	0	0	0	0

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within Dept of Health Specifications
No adjustments or changes have been made to this instrument

INSPECTING OFFICER

SIGNATURE	PRINT FULL NAME
Melvin Taber	Melvin Taber
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
230 211 10/02/15	660 335 6823

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Blvd.
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 13002
EXPIRATION DATE: June 19, 2015 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **13002** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1217** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **June 20, 2013** The expiration date for this lot number is **June 19, 2015** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

2

PERMIT

TYPE II

MELVIN L TABER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/2/2013

NUMBER 230211

EXPIRES 10/2/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI

DEPARTMENT OF HEALTH AND SENIOR SERVICES

BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TABER, MELVIN

Permit No 230211

Date Issued 10/2/2013 **Date Expires** 10/2/2015

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097
04/15/15

ARREST TIME: 13:00
SUBJECT NAME:
TABER/M/L
DOB: 01/23/53 SEX: M
STATE/D.L.: MO/123456789
ARRESTING OFFICER:
TABER/MELVIN/L
OFFICER I.D.: 902
TESTING OFFICER:
TABER/MELVIN/L
OFFICER I.D.: 902
PERMIT NUMBER: 230211
EXPIRATION DATE: 10/02/15
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE REPORT
RADIO INTERFERENCE TEST

-- BREATH ANALYSIS --

BLANK TEST	.000	13:52
INTERNAL STANDARD	VERIFIED	13:52
RADIO INTERFERENCE		

erator Signature _____

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097
04/15/15
13:40

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

Operator Signature _____

2208-02

Face This Side Down – This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
SWEET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097
04/15/15

TESTING OFFICER:
TADER/MELVIN/L
OFFICER I.D.: 902
PERMIT NUMBER: 230211
EXPIRATION DATE: 10/02/15
MISCELLANEOUS DATA:
MONTHLY MAINTENANCE TEST
SUPERVISOR MODE

SUPERVISOR MODE

BLANK TEST	.000	13:43
INTERNAL STANDARD	VERIFIED	13:43
EXTERNAL STANDARD	.099	13:44
BLANK TEST	.000	13:44
EXTERNAL STANDARD	.100	13:45
BLANK TEST	.000	13:45
EXTERNAL STANDARD	.099	13:46
BLANK TEST	.000	13:46

N = 3
SIM. = .099
AVG. = .0993

Operator Signature _____

2208-02