



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED** REPORT #6  
 By Carol Day at 12:35 pm, Jul 13, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>940097</b>	NAME OF AGENCY <b>Sweet Springs P.D.</b>	DATE OF INSPECTION <b>06/28/15</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>324 S. Miller Sweet Springs</b>		TIME OF INSPECTION <b>21:57</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>06/28/15 21:57</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>47</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Rep Co Marketing Inc** LOT # **15001** EXP. DATE **05/20/2012**

SIMULATOR TEMP (34°C ± 0.2°C) **34°C ± 0.2°C** °C SIMULATOR SN **SD3144** EXP. DATE **08/25/15**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <b>.101%</b>	TEST 2 <b>.101%</b>	TEST 3 <b>.101%</b>
---------------------	---------------------	---------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED) **Radio Interference Detected.**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>0</b>	(0-.04) <b>0</b>	(.05-.09) <b>1</b>	(.10-.14) <b>0</b>	(.15-.19) <b>1</b>	OVER .19 <b>0</b>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

*This instrument is operating within Dept. of Health Specifications using .100% solution from Rep Co marketing bearing lot # 15001. No adjustments or changes were made.*

**INSPECTING OFFICER**

SIGNATURE <i>Melvin Taber</i>	PRINT FULL NAME <b>Melvin Taber</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>230211 10/02/2015</b>	TELEPHONE NUMBER <b>660 335 6823</b>

RETURN COMPLETED REPORT TO THE:  
 Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 15001**

**EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MELVIN L TABER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/2/2013

NUMBER 230211

EXPIRES 10/2/2015

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

,acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator TABER, MELVIN  
 Permit No 230211  
 Date Issued 10/2/2013 Date Expires 10/2/2015

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
SWEET SPRINGS POLICE DEPARTMENT

DAC DATAMASTER SERIAL NUMBER 940007  
06-28-15

TESTING OFFICER:  
TAYLOR, MELVIN L  
OFFICER I.D.: 002  
PERMIT NUMBER: 230211  
EXPIRATION DATE: 10-02-15  
MISCELLANEOUS DATA:  
MONTHLY MAINTENANCE  
SUPERVISOR MODE

SUPERVISOR MODE

BLANK TEST	.000	22:01
INTERNAL STANDARD	VERIFIED	22:01
EXTERNAL STANDARD	.101	22:01
BLANK TEST	.000	22:02
EXTERNAL STANDARD	.101	22:02
BLANK TEST	.000	22:03
EXTERNAL STANDARD	.101	22:03
BLANK TEST	.000	22:04

N = 3  
SIM. = .899  
MS. = .101

Operator Signature

2208-02

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
SWEET SPRINGS POLICE DEPARTMENT

DAC DATAMASTER SERIAL NUMBER 940007  
06-28-15  
21:07

DIAGNOSTIC CHECK

COMPUTER:	OKAY
PROGRAM (04-07-2009):	OKAY
HEATERS	
SAMPLE CHAMBER:	45c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTED TEST

! " # \$ % ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z I \ ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ \*

Operator Signature

2208-02

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CAMEL SPRINGS POLICE DEPARTMENT

INC DATAMASTER SERIAL NUMBER 940097  
06/20/15

ARREST TIME: 21:00

SUBJECT NAME:

TRUER/M/L

DOB: 01-28-50 SEX: M

STATE I.D. NO: 123456789

ARRESTING OFFICER:

TRUER/MELVIN-L

OFFICER I.D. NO: 902

TESTING OFFICER:

TRUER/MELVIN-L

OFFICER I.D. NO: 902

PERMIT NUMBER: 200211

EXPIRATION DATE: 12/31/15

MISCELLANEOUS DATA:

MONTHLY MAINTENANCE

RADIO INTERFERENCE TEST

## BREATH ANALYSIS

BLANK TEST	✓	00:00
INTERNAL STANDARD	VERIFIED	00:00
RADIO INTERFERENCE		

Operator Signature

