



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

**RECEIVED**  
By Brian Lutmer at 1:21 pm, Oct 14, 2015

REPORT #8

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <u>940097</u>	NAME OF AGENCY <u>Sweet Springs RD</u>	DATE OF INSPECTION <u>10/01/15</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>324 S. Miller Sweet Springs</u>		TIME OF INSPECTION <u>21:20</u>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <u>10/01/15 21:20</u>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <u>49<sup>e</sup></u> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <u>Repro marketing</u> LOT # <u>15001</u> EXP. DATE <u>05/20/17</u>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <u>34°C ± 0.2</u> °C SIMULATOR SN <u>SD3144</u> EXP. DATE <u>08/25/15</u>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <u>100%</u>	TEST 2 <u>100%</u>	TEST 3 <u>100%</u>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED) RFI detected

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	OVER .19 <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

This instrument is operating within Dept of Health specifications using 100% solution from repro marketing bearing lot # 15001 and expiration date 05/20/17

INSPECTING OFFICER	
SIGNATURE <u>Melvin Taber</u>	PRINT FULL NAME <u>Melvin Taber</u>
TYPE, PERMIT NUMBER/EXPIRATION DATE <u>230211 10/02/15</u>	TELEPHONE NUMBER <u>660 335 6823</u>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63901**

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 15001**

**EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**MELVIN L TABER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/2/2013

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 230211

acting director  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 10/2/2015

MO 680-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator **TABER, MELVIN**  
 Permit No **230211**  
 Date issued **10/2/2013** Date Expires **10/2/2015**

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CUMMET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097  
10-01-15

ARREST TIME: 10:50

SUBJECT NAME:

TABER, MELVIN L.

DOB: 01/29/58 SEX: M

STATE I.D.# MO: 102456789

ARRESTING OFFICER:

TABER, MELVIN L.

OFFICER I.D.# 988

TESTING OFFICER:

TABER, MELVIN L.

OFFICER I.D.# 988

PERMIT NUMBER: 200211

EXPIRATION DATE: 10-02-15

MISCELLANEOUS DATA:

MONTHLY MAINTENANCE REPORT

RADIO INTERFERENCE TEST

### BREATH ANALYSIS

BLANK TEST	.000	21:30
INTERNAL STANDARD	VERIFIED	21:31
RADIO INTERFERENCE		

Operator Signature \_\_\_\_\_

2208-02

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CUMMET SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 940097  
10-01-15  
21:29

### BIRD-NOSTIC CHECK

COMPUTER:	OKAY
PROGRAM (04-07-2007):	OKAY
HEATERS:	
SAMPLE CHAMBER:	49L
FLOW DETECTOR:	OKAY
PUMP:	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
DURATE STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

10/01/15 10:50:00  
MELVIN L. TABER  
01/29/58 M

Operator Signature \_\_\_\_\_

2208-02

Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
COURT SPRINGS POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 948897  
10/01/15

TESTING OFFICER:  
TERRY MELVIN  
OFFICER I.D.# 388  
PERMIT NUMBER: 258311  
EXPIRATION DATE: 10/02/18  
MISCELLANEOUS DATA:  
RARELY MAINTENANCE TEST  
SUPERVISOR MODE

SUPERVISOR MODE

BLANK TEST	.000	21.02
INTERNAL STANDARD	VERIFIED	21.00
EXTERNAL STANDARD	.100	21.01
BLANK TEST	.000	21.24
EXTERNAL STANDARD	.100	21.00
BLANK TEST	.000	21.05
EXTERNAL STANDARD	.100	21.01
BLANK TEST	.000	21.04

N = 3  
SIN. = .009  
AVG. = .1

Operator Signature 