



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 6:32 am, Feb 03, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 940228	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 01/27/2015
LOCATION OF INSTRUMENT (STREET AND CITY) Monroe City Police Department, 300 North Main Street, Monroe City, MO		TIME OF INSPECTION 11:28 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/27/2015 11:28
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Guth Laboratories** LOT # **13290** EXP. DATE **10/29/2015**

SIMULATOR TEMP (34°C ± 0.2°C) **34.0** °C SIMULATOR SN **MP2456** EXP. DATE **09/09/2015**

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .096	TEST 2 .097	TEST 3 .098
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(.0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	----------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME James Johnson, Trooper
TYPE II PERMIT NUMBER/EXPIRATION DATE 240173 04/22/2016	TELEPHONE NUMBER (660) 385-2132

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**



GUTH LABORATORIES, INC.

890 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **13290** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 31, 2013**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1202%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 29, 2015** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

**Ted L. Pauley, President
GUTH LABORATORIES, INC.**

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JAMES P JOHNSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2014

NUMBER 240173

EXPIRES 4/22/2016

580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator JOHNSON, JAMES
Permit No 240173
Date Issued 4/22/2014 Date Expires 4/22/2016

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 246008
01/27/11
11:00

DIAGNOSTIC CHECK

COMPUTER: OKAY

PROGRAM (04-07 2003): OKAY

HEADERS
SAMPLE CHAMBER: OKAY

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY


QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / : ; < = > ? @ [\] ^ _ ` { | } ~
0123456789 : ; < = > ? @ [\] ^ _ ` { | } ~
var of printer 11-11

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket


MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 246008
01/27/11

HOURS TIME: 11:00
SUBJECT NUMBER:
RFI TEST
DOB: 11/11/11 SEX: M
STATE D.L.I.: MO-NA
ARRESTING OFFICER:
IA
OFFICER I.D.: NA
TESTING OFFICER:
JOHNSON, JAMES P
OFFICER I.D.: 407
PERMIT NUMBER: 246172
EXPIRATION DATE: 04/20/14
TESTED LABORATORY:
RADIO FREQUENCY INTERFERENCE TEST

BREATH ANALYSIS

BLANK TEST	1.000	1.000
INTERNAL STANDARD	0.000	1.000
RADIO INTERFERENCE		

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER: 240226
01-07-15

TESTING OFFICER:
LOWINSON, JAMES P
OFFICER I.D. #: 487
PERMIT NUMBER: 240173
EXPIRATION DATE: 04-08-15
MISCELLANEOUS DATA:

SUPERVISION MODE

BLANK TEST	.000	11:04
INTERNAL STANDARD	VERIFIED	11:04
EXTERNAL STANDARD	.000	11:06
BLANK TEST	.000	11:07
EXTERNAL STANDARD	.007	11:09
BLANK TEST	.000	11:04
EXTERNAL STANDARD	.009	11:06
BLANK TEST	.000	11:07

N = 3
S.D. = .1
A.V.G. = .007

Operator Signature _____

