



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

RECEIVED
 By Brian Lutmer at 4:26 pm, Jan 11, 2015

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950068	NAME OF AGENCY MEXICO PUBLIC SAFETY DEPARTMENT	DATE OF INSPECTION 01/06/2015
LOCATION OF INSTRUMENT (STREET AND CITY) 300 NORTH COAL STREET, MEXICO, MO 65265		TIME OF INSPECTION 2:50 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 01/06/15 2:50pm
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER _____ 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER REPCO MARKETING LOT # 14001 EXP. DATE 04/30/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34 °C SIMULATOR SN SD 2669 EXP. DATE 04/08/2015

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .100	TEST 3 .101
--------------	--------------	--------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE 	PRINT FULL NAME STUART J. THOMPSON
TYPE II PERMIT NUMBER/EXPIRATION DATE 230078 04/26/2015	TELEPHONE NUMBER (573) 581-2503

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901

COPY

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MEXICOD DEPARTMENT OF PUBLIC SAFETY
390 DETRMASTER SERIAL NUMBER 950060
01/05/15
14150

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM (06-07-2009): OKAY
HERTERS
SAMPLE CHAMBER: 49C
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
ORLIZATION: OKAY
PRINTER TEST
I#\#\#\#()***-,/0123456789:;(<=>?@ABDEFG
HIJKLMNORSTUVWXYZ[\]^_`abcdefg|hijklmno
pqrstuvwxyz{|}**

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
MEXICOD DEPARTMENT OF PUBLIC SAFETY
390 DETRMASTER SERIAL NUMBER 950060
01/05/15

TESTING OFFICER:

THOMPSON/STUBERT/J
OFFICER I,2,1 14
PERMIT NUMBER: 200079
EXPIRATION DATE: 04/24/15
MISCELLANEOUS DATA:
JPN: TEST

--- SUPERVISION MODE ---

BLANK TEST .000 14150
INTERNAL STANDARD .000 14150
EXTERNAL STANDARD .101 14150
BLANK TEST .000 14150
EXTERNAL STANDARD .100 14150
BLANK TEST .000 14150
EXTERNAL STANDARD .101 14150
BLANK TEST .000 14150
N = 3
STD. = .1
RMS. = .1005

Operator Signature



Face This Side Down - This Edge In First

BAC DataMaster
Evidence Ticket

STATE OF MISSOURI
VENUE DEPARTMENT OF PUBLIC SAFETY
MCD DATAMASTER SERIAL NUMBER 04000
01/05/15

COPY

ARREST TIME: 14:00
SUBJECT NAME:

PC1/TEST

DOB: 01/21/91 SEX: M

STATE/CLL: MO/12345

ARRESTING OFFICER:

THOMPSON/STUART/J

OFFICER I.D.# 14

TESTING OFFICER:

THOMPSON/STUART/J

OFFICER I.D.# 14

PERMIT NUMBER: 220070

EXPIRATION DATE: 04/26/15

MISCELLANEOUS DATA:

QTY TEST

----- BREATH ANALYSIS -----

BLANK TEST : 0000 14:55
INTERNAL STANDARD : 0000 14:55
PC010 INTERFERENCE : 0000 14:55

Handwritten marks and scribbles at the bottom of the page.

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

COPY



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

**PERMIT
 TYPE II**

STUART J THOMPSON III

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field services and repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 04/26/2013

W.S. [Signature]

NUMBER 230078

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

EXPIRES 04/26/2015

W.S. [Signature]

EXP 04/26/2015

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

COPY