



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

**RECEIVED**  
 By Carol Day at 9:07 am, Jul 02, 2015

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days) or when the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN <b>950090</b>	NAME OF AGENCY <b>Pheeps Co. Sheriff's Dept.</b>	DATE OF INSPECTION <b>6-25-15</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>500 W. 2<sup>ND</sup> Rolla, Mo 65401</b>		TIME OF INSPECTION <b>2033</b>

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) <b>6-25-15 2033</b>
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER <b>49</b> °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

<input checked="" type="checkbox"/> INDICATOR LIGHTS
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER <b>Guth Labs</b> LOT # <b>14220</b> EXP. DATE <b>9-24-16</b>
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) <b>34</b> °C SIMULATOR SN <b>SD 1313</b> EXP. DATE <b>1-21-16</b>

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 <b>.097</b>	TEST 2 <b>.097</b>	TEST 3 <b>.097</b>
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <b>~</b>	(0-.04) <b>—</b>	(.05-.09) <b>1</b>	(.10-.14) <b>1</b>	(.15-.19) <b>—</b>	OVER .19 <b>—</b>
-------------------	------------------	--------------------	--------------------	--------------------	-------------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).  
**N/A**

<b>INSPECTING OFFICER</b>	
SIGNATURE <b>[Signature]</b> 204	PRINT FULL NAME <b>Paul J. Lambert</b>
TYPE II PERMIT NUMBER/EXPIRATION DATE <b>240292 6-27-16</b>	TELEPHONE NUMBER <b>573-308-6213</b>

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901**

Face This Side Down - This Edge In First

**BAC DataMaster**  
**Evidence Ticket**

STATE OF MISSOURI  
PHELPS COUNTY SHERIFFS DEPT

BAC DATAMASTER SERIAL NUMBER 950090  
06/25/15

TESTING OFFICER:  
LAMBERT/PAUL/J  
OFFICER I.D.# 204  
PERMIT NUMBER: 248202  
EXPIRATION DATE: 06/27/16  
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	20:26
INTERNAL STANDARD	VERIFIED	20:26
EXTERNAL STANDARD	.097	20:37
BLANK TEST	.000	20:37
EXTERNAL STANDARD	.097	20:38
BLANK TEST	.000	20:39
EXTERNAL STANDARD	.097	20:39
BLANK TEST	.000	20:40

N = 3  
SIM. = .1  
R/S. = .097

Operator Signature



2208-02

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
CHELSEA COUNTY SHERIFFS DEPT

BAC DATAMASTER SERIAL NUMBER 250090  
06-05-15  
20:30

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM (04-07-2003):	OKAY
HEATERS	
SAMPLE CHAMBER:	40c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;@<=>?`ABCD EFG  
HIJKLMN O PQRSTU VWXYZ[\]^\_`ab cdefghijklmno  
pqrstuvwxyz{|}~\*

208-02

Operator Signature \_\_\_\_\_



**BAC DataMaster  
Evidence Ticket**

STATE OF MISSOURI  
WHEELER COUNTY SHERIFFS DEPT

BAC DATAMASTER SERIAL NUMBER 950090  
06/25/15

ARREST TIME: 20:00  
SUBJECT NAME:  
DOOFUS/BOB/7  
DOB: 02/02/82 SEX: M  
STATE/D.L.: MO/L190298  
ARRESTING OFFICER:  
LAMBERT/PAUL/J  
OFFICER I.D.: 204  
TESTING OFFICER:  
LAMBERT/PAUL/J  
OFFICER I.D.: 204  
PERMIT NUMBER: 249092  
EXPIRATION DATE: 06/27/15  
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

RADIO INTERFERENCE

Operator Signature

*Paul Lambert*  
204

2208-02

2208-02



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**PAUL J LAMBERT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/27/2014

NUMBER 240292

EXPIRES 6/27/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator LAMBERT, PAUL  
 Permit No 240292  
 Date Issued 6/27/2014 Date Expires 6/27/2016



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*