



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
 BREATH ALCOHOL PROGRAM
DATAMASTER MAINTENANCE REPORT

REPORT #6

RECEIVED

By Carol Day at 2:29 pm, Jan 11, 2016

Complete this report at the time of the regular monthly preventive maintenance (do not exceed 35 days).
 Complete this report whenever the instrument is serviced or repaired and retained for 30 days.
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950090	NAME OF AGENCY Pheps Co. Sheriff's Dept.	DATE OF INSPECTION 12-31-15
LOCATION OF INSTRUMENT (STREET AND CITY) 500 W. 2ND, Rolla, Mo 65401		TIME OF INSPECTION 0614

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 12-31-15 0614
<input type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER **Guth Labs** LOT # **14220** EXP. DATE **9-24-16**

SIMULATOR TEMP (34°C ± 0.2°C) **34** °C SIMULATOR SN **SD1313** EXP. DATE **1-21-16**

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 .097	TEST 2 .098	TEST 3 .099
--------------------	--------------------	--------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	OVER .19
----------	---------	-----------	-----------	-----------	----------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

INSPECTING OFFICER	
SIGNATURE 	PRINT FULL NAME Paul J. Lambert
TYPE II PERMIT NUMBER EXPIRATION DATE 250072 3-31-17	TELEPHONE NUMBER 573-426-3860

RETURN COMPLETED REPORT TO THE: **Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Blvd.
 Poplar Bluff, MO 63901**

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
PHILIP COUNTY SURVIVOR BODY

BAC DATAMASTER SERIAL NUMBER 950000
12/31/15

TESTING OFFICER:
LANDERT/PAU/J
OFFICER I.D.# 294
PERMIT NUMBER: 250072
EXPIRATION DATE: 03/31/17
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	00:10
INTERNAL STANDARD	VERIFIED	00:10
EXTERNAL STANDARD	.007	00:10
BLANK TEST	.009	00:10
EXTERNAL STANDARD	.002	00:20
BLANK TEST	.000	00:20
EXTERNAL STANDARD	.000	00:21
BLANK TEST	.000	00:22

N = 0
S.M. = .1
A.V.G. = .000

Operator Signature _____

DL/204

2208-02

BAC DataMaster Evidence Ticket

STATE OF MISSOURI
SHELPS COUNTY SHERIFFS DEPT


DAD DATAMASTER SERIAL NUMBER 250000
12/01/15

ARREST TIME: 08:00
SUBJECT NAME:
DOE, JIM/D
DOB: 02/02/02 SEX: M
STATE/D.L.L.# MO/11000000
REQUESTING OFFICER:
LAMBERT, PAUL/J
OFFICER I.D.# 204
TESTING OFFICER:
LAMBERT, PAUL/J
OFFICER I.D.# 204
PERMIT NUMBER: 250070
EXPIRATION DATE: 03/31/17
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.002	06:27
INTERNAL STANDARD	VERIFIED	06:27
RADIO INTERFERENCE		

Operator Signature _____





STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

PAUL J LAMBERT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/31/2015

NUMBER 250072

EXPIRES 3/31/2017

[Signature]
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 539-9-1 (6-13)

LAB-4 (06-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **LAMBERT, PAUL**
 Permit No **250072**
 Date Issued **3/31/2015** Date Expires **3/31/2017**



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14220 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 25, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is September 24, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.