

**RECEIVED**

By Ellen Strawsine at 2:11 pm, Oct 16, 2015



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM  
**DATAMASTER MAINTENANCE REPORT**

REPORT #6

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
 Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
 Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 950124	NAME OF AGENCY Kirksville Police Department	DATE OF INSPECTION 10/02/2015
-------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 119 E McPherson, Kirksville, MO, 63501	TIME OF INSPECTION 6:44 pm
--	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIAGNOSTIC CHECK (PRINTOUT ATTACHED)      DATE AND TIME (from printout) 10/02/2015 18:44 hours

COMPUTER       DETECTOR

PROGRAM       FILTERS

HEATERS SAMPLE CHAMBER 49 °C       QUARTZ STANDARD

FLOW DETECTOR       CALIBRATION

PUMP HIGH SPEED       PRINTER

INDICATOR LIGHTS

SIMULATOR SOLUTION SUPPLIER Guth Labs      LOT # 14200      EXP. DATE 08/05/2016

SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C      SIMULATOR SN SD2271      EXP. DATE 11/24/2015

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 $\blacktriangleleft$ .100	TEST 2 $\blacktriangleleft$ .101	TEST 3 $\blacktriangleleft$ .101
----------------------------------	----------------------------------	----------------------------------

PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	OVER .19	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	----------	---

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

**INSPECTING OFFICER**

SIGNATURE 	PRINT FULL NAME Juan B Chairez
---------------	-----------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 250152      07/22/2017	TELEPHONE NUMBER (660) 785-6945
---	------------------------------------

RETURN COMPLETED REPORT TO THE:      Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Blvd.  
 Poplar Bluff, MO 63901



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**


**JUAN B CHAIREZ**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):


**DATAMASTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

  
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 250152

  
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 7/22/2017

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator CHAIREZ JUAN  
 Permit No 250152  
 Date Issued 7/22/2015 Date Expires 7/22/2017



**GUTH LABORATORIES, INC.**

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 14200 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 6, 2014, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1213% (w/vol) ethyl alcohol. The expiration date for this lot number is August 5, 2016 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN122211-02 whose values are traceable to NIST*

*All balances are calibrated annually by an outside agency using NIST traceable weights  
Calibration verification is done prior to each use utilizing NIST traceable weights*



Face This Side Down - This Edge In First

# BAC DataMaster Evidence Ticket

STATE OF MISSOURI  
KIRKSVILLE POLICE DEPARTMENT

BAC DATAMASTER SERIAL NUMBER 908124  
18 40/15

TESTING OFFICER:  
CHRISTY JUAN-C  
OFFICER I.D.# 360  
PERMIT NUMBER: 258103  
EXPIRATION DATE: 07-02-17  
MISCELLANEOUS DATA:

SUPERVISOR NAME

BLANK TEST	.000	18:40
INTERNAL STANDARD	VERIFIED	18:40
EXTERNAL STANDARD	.100	18:40
BLANK TEST	.000	18:40
INTERNAL STANDARD	.101	18:40
BLANK TEST	.000	18:40
EXTERNAL STANDARD	.101	18:40
BLANK TEST	.000	18:40

N = 7  
S.D. = .1  
A.V.C. = .1000

Operator Signature J. G. KPO387