



AS IV Serial no: 026999  
Version no: 7409

TEST RECORD 01790

Temp Date Time 210L s/  
Air Blank: 03/05/16 13:45 .000  
Calibration Check:  
21 03/05/16 13:45 .103

Subject Name

Subject I.D.

Operator Name, I.D.

ROBERT PAUL 250171  
Location

AS IV Serial no: 026999  
Version no: 7409

TEST RECORD 01791

Temp Date Time 210L s/  
Air Blank: 03/05/16 13:48 .000  
Calibration Check:  
22 03/05/16 13:48 .099

Subject Name

Subject I.D.

Operator Name, I.D.

ROBERT PAUL 250171  
Location

AS IV Serial no: 026999  
Version no: 7409

TEST RECORD 01792

Temp Date Time 210L s/  
Air Blank: 03/05/16 13:50 .000  
Calibration Check:  
24 03/05/16 13:50 .097

Subject Name

Subject I.D.

Operator Name, I.D.

ROBERT PAUL 250171  
Location

AS IV Serial no: 026999  
Version no: 7409

TEST RECORD 01793

Temp Date Time 210L s/  
Air Blank: 03/05/16 13:52

Subject Name

Subject I.D.

Operator Name, I.D.

ROBERT PAUL 250171  
Location



**Airgas USA LLC (LAB)**  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

**Certificate of Analysis**

**Test Date:** 8-Oct-2014

**Customer Name**  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Lot # AG428002**

<b>Exp. Date</b> 7-Oct-2016	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm) Balance
--------------------------------	-------------------------	---	--

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2014.10.08 12:15:00 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

**Analyst:** Rod Marsala  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**ROBERT L PAUL**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/8/2015

NUMBER 250131

EXPIRES 6/8/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB 4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator PAUL, ROBERT  
 Permit No 250131  
 Date Issued 6/8/2015 Date Expires 6/8/2017