



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Carol Day at 10:16 am, Jun 09, 2016

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030788	PRINTER SN 093.3567.011	DATE OF INSPECTION 05/31/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 400 First Street, Hillsboro, MO 63050		TIME OF INSPECTION 1:01 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Lab LOT # 15120 EXP. DATE 04/29/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2222 SIMULATOR EXP DATE 04/13/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .104

TEST 2 ➡ .103

TEST 3 ➡ .103

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	(.05-.09)	1	(.10-.14)	1	(.15-.19)	(OVER .19)	1
----------	---	---------	-----------	---	-----------	---	-----------	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument operates within the guidelines set forth by the Department of Health and Senior Services.

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
 Deputy Nick Gamm #549

TYPE II PERMIT NUMBER/EXPIRATION DATE  
 260216 / 05/03/2018

TELEPHONE NUMBER  
 (636) 797-5000

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 030788  
Version no: 532B

TEST RECORD 00182

Temp	Date	Time	s/	210L
-----				
Air Blank:				
	05/31/16	13:18	.000	
Calibration Check:				
	26	05/31/16	13:18	.104

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Dep Nick Gamm 549

Location

400 First St

Hillsboro, MO 63050

AS IV Serial no: 030788  
Version no: 532B

TEST RECORD 00183

Temp	Date	Time	s/	210L
-----				
Air Blank:				
	05/31/16	13:22	.000	
Calibration Check:				
	27	05/31/16	13:22	.103

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Dep Nick Gamm 549

Location

400 First St

Hillsboro, MO 63050

AS IV Serial no: 030788  
Version no: 532B

TEST RECORD 00184

Temp	Date	Time	s/	210L
-----				
Air Blank:				
	05/31/16	13:25	.000	
Calibration Check:				
	27	05/31/16	13:25	.103

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Dep Nick Gamm 549

Location

400 First St

Hillsboro, MO 63050

AS IV Serial no: 030788  
Version no: 532B

TEST RECORD 00185

Temp	Date	Time	s/	210L
-----				
VOID: RFI				
	12	05/31/16	13:27	

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Dep Nick Gamm 549

Location

400 First St.

Hillsboro, MO 63050



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**NICHOLAS J GAMM**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/3/2016

NUMBER 260216

EXPIRES 5/3/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **GAMM, NICHOLAS**  
Permit No **260216**  
Date Issued **5/3/2016**    Date Expires **5/3/2018**



**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

## CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

***NIST Traceability:***

*Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*