

AS IV Serial no: 030789
Version no: 532B

TEST RECORD 00096

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
05/31/16 13:14 .000
Calibration Check:
23 05/31/16 13:14 .099

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Dep. Nick Gramm 549

Location

400 First St.

Hillsboro, MO 63050

AS IV Serial no: 030789
Version no: 532B

TEST RECORD 00097

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
05/31/16 13:16 .000
Calibration Check:
24 05/31/16 13:16 .101

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Dep. Nick Gramm 549

Location

400 First St.

Hillsboro, MO 63050

AS IV Serial no: 030789
Version no: 532B

TEST RECORD 00098

Temp	Date	Time	s/	210L
------	------	------	----	------

Air Blank:
05/31/16 13:19 .000
Calibration Check:
25 05/31/16 13:19 .100

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Dep. Nick Gramm 549

Location

400 First St.

Hillsboro, MO 63050

AS IV Serial no: 030789
Version no: 532B

TEST RECORD 00099

Temp	Date	Time	s/	210L
------	------	------	----	------

VOID: RFI
12 05/31/16 13:22

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Dep. Nick Gramm 549

Location

400 First St.

Hillsboro, MO 63050



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
NICHOLAS J GAMM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/3/2016

NUMBER 260216

EXPIRES 5/3/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **GAMM, NICHOLAS**
 Permit No **260216**
 Date issued **5/3/2016** Date Expires **5/3/2018**



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.