

RECEIVED

By Brian Lutmer at 2:40 pm, Jan 29, 2016



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030794 PRINTER SN 84.9324-031 DATE OF INSPECTION 1/19/16

LOCATION OF INSTRUMENT (STREET AND CITY) 1781 Zumbel Rd., St Charles, Mo 63303 TIME OF INSPECTION 1514hrs

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) ok
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C) ok
- PRINTER WORKING PROPERLY ok
- TIME AND DATE DISPLAYING PROPERLY ok

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS-MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 15120 EXP. DATE 4/29/17

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0°C SIMULATOR SN SD2248 SIMULATOR EXP DATE 1/4/17

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .103 TEST 2 • .102 TEST 3 • .102

RFI DETECTOR OPERATING ok

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	<u>2</u>	(0-.04)	<u>0</u>	(.05-.09)	<u>0</u>	(.10-.14)	<u>0</u>	(.15-.19)	<u>1</u>	(OVER .19)	<u>0</u>
----------	----------	---------	----------	-----------	----------	-----------	----------	-----------	----------	------------	----------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE Daniel Gibbons 324 PRINT NAME Daniel Gibbons 324

TYPE II PERMIT NUMBER/EXPIRATION DATE 250021 1/20/2017 TELEPHONE NUMBER (636) 949-3300

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 030794
Version no: 004C

TEST RECORD 01168

Temp Date Time 210L ^{s/}

Air Blank:
01/19/16 15:14 .000
Calibration Check:
21 01/19/16 15:14 .103

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbeth Rd.

St Charles, Mo 63303

AS IV Serial no: 030794
Version no: 004C

TEST RECORD 01169

Temp Date Time 210L ^{s/}

Air Blank:
01/19/16 15:16 .000
Calibration Check:
22 01/19/16 15:16 .102

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbeth Rd

St Charles, Mo 63303

AS IV Serial no: 030794
Version no: 004C

TEST RECORD 01170

Temp Date Time 210L ^{s/}

Air Blank:
01/19/16 15:17 .000
Calibration Check:
22 01/19/16 15:17 .102

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbeth Rd.

St Charles, Mo 63303

AS IV Serial no: 030794
Version no: 004C

TEST RECORD 01171

Temp Date Time 210L ^{s/}

Uoid: RFI
12 01/19/16 15:18

Subject Name

Subject I.D.

Operator Name, I.D.

D. Gibbons 324

Location

1781 Zumbeth Rd

St Charles, Mo 63303