



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 035485	PRINTER SN 096.3580.926	DATE OF INSPECTION 02/10/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 101 Sheriff Dierker Court, O'Fallon		TIME OF INSPECTION 11:32 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)	PASSED
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)	23°C
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY	PASSED
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY	02/10/2016 11:32 hrs.

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters, Inc. LOT # AG525303 EXP. DATE 09/10/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE  
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE  
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .080      TEST 2 • .080      TEST 3 • .079

RFI DETECTOR OPERATING PASSED

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME Officer Fournell, D. DSN 570
---------------	--

TYPE II PERMIT NUMBER/EXPIRATION DATE 240274 - 06/13/2016	TELEPHONE NUMBER (636) 949-3000
--	------------------------------------

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**DAVID FOURNELL**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/13/2014

NUMBER 240274

EXPIRES 6/13/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator FURNELL, DAVID  
 Permit No 240274  
 Date Issued 6/13/2014 Date Expires 6/13/2016

AS IV Serial no: 635485  
Version no: 532B

TEST RECORD 06271 s/  
Temp Date Time 210L

Air Blank:  
02/10/16 11:32 .000  
Calibration Check:  
23 02/10/16 11:32 .000

Subject Name

Subject I.D.

Operator Name, I.D.

Fournell 240274  
Location

101 SHERIFF DRIVE CT

O'FALLON, MD

AS IV Serial no: 635485  
Version no: 532B

TEST RECORD 06272 s/  
Temp Date Time 210L

Air Blank:  
02/10/16 11:34 .000  
Calibration Check:  
24 02/10/16 11:34 .000

Subject Name

Subject I.D.

Operator Name, I.D.

Fournell 240274  
Location

101 SHERIFF DRIVE CT

O'FALLON MD

AS IV Serial no: 635485  
Version no: 532B

TEST RECORD 06273 s/  
Temp Date Time 210L

Air Blank:  
02/10/16 11:36 .000  
Calibration Check:  
25 02/10/16 11:36 .079

Subject Name

Subject I.D.

Operator Name, I.D.

Fournell 240274  
Location

101 SHERIFF DRIVE CT

O'FALLON MD

AS IV Serial no: 635485  
Version no: 532B

TEST RECORD 06274 s/  
Temp Date Time 210L

UID: RFI  
12 02/10/16 11:38

Subject Name

Subject I.D.

Operator Name, I.D.

Fournell 240274  
Location

101 SHERIFF DRIVE CT

O'FALLON MD