



\*om<sup>2</sup> \*

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 00523

Temp Date Time 210L  
-----  
s/

Air Blank:  
07/05/16 09:31 .000  
Calibration Check:  
26 07/05/16 09:31 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

\*=3<sup>u</sup> \*

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 00524

Temp Date Time 210L  
-----  
s/

VOID: RFI  
12 07/05/16 09:32

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 00521

Temp Date Time 210L  
-----  
s/

Air Blank:  
07/05/16 09:27 .000  
Calibration Check:  
26 07/05/16 09:27 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location

\*=3<sup>u</sup> \*

AS IV Serial no: 062086  
Version no: 532B

TEST RECORD 00522

Temp Date Time 210L  
-----  
s/

Air Blank:  
07/05/16 09:29 .000  
Calibration Check:  
26 07/05/16 09:29 .100

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 15-Sep-2015

**Lot # AG525701 Model 108caccd**


<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
14-Sep-2017	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2015.09.15 16:04:10 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Analyst:   
 Rod Marsala

*ISO 17025:2005 A2LA accredited. Certificate Number 2989.01*



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**MICHAEL RICHARDSON**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/31/2016

NUMBER 260168

EXPIRES 3/31/2018

\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

\_\_\_\_\_  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator RICHARDSON, MICHAEL  
Permit No 260168  
Date Issued 3/31/2016 Date Expires 3/31/2018