



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 062093	PRINTER SN 093.6563.014	DATE OF INSPECTION 01/30/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 1001 Heroes Way, Sugar Creek		TIME OF INSPECTION 5:02 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 15050 EXP. DATE 03/09/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3136 SIMULATOR EXP DATE 07/10/2016

CALIBRATION CHECK— (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ✎ .102

TEST 2 ✎ .103

TEST 3 ✎ .102

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	3	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Wesley H. Edwards</i>	PRINT NAME Wesley H. Edwards
---------------------------------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 250213 / 10/01/2017	TELEPHONE NUMBER (816) 252-7058
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-634-5170

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15050 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 11, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is March 9, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 602693
Version no: 5324

TEST RECORD 06494

Temp Date Time ^{w/} 2100

Air Blank
01/30/16 17:02 .609
Calibration Check
02 01/30/16 17:02 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO

AS IV Serial no: 602693
Version no: 5324

TEST RECORD 06496

Temp Date Time ^{w/} 2100

Air Blank
01/30/16 17:00 .609
Calibration Check
24 01/30/16 17:00 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO

AS IV Serial no: 602693
Version no: 5324

TEST RECORD 06495

Temp Date Time ^{w/} 2100

Air Blank
01/30/16 17:04 .609
Calibration Check
23 01/30/16 17:04 .103

Subject Name

Subject I.D.

Operator Name, I.D.

Edwards 908

Location

1001 Heroes Way

Sugar Creek, MO

AS IV Serial no: 602693
Version no: 5324

TEST RECORD 06497

Temp Date Time ^{w/} 2100

UNIT: RFI
12 01/30/16 17:00

Subject Name

Subject I.D.

Edwards 908

Operator Name, I.D.

Location

1001 Heroes Way

Sugar Creek, MO



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

STEPHEN C MYERS

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/1/2015

NUMBER 250210

EXPIRES 10/1/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-1 (RS-10)

MO 580-0771 (6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **MYERS, STEPHEN**
Permit No **250210**
Date Issued **10/1/2015** Date Expires **10/1/2017**