





STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**BRETT J DAVIS**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOXILYZER 5000, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/18/2015

NUMBER 250183

EXPIRES 8/18/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator **DAVIS, BRETT**  
 Permit No **250183**  
 Date Issued **8/18/2015** Date Expires **8/18/2017**



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 10-Feb-2015

**Lot #** AG504002

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
9-Oct-2016	30	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

**Certification Traceable to N.I.S.T. RGM Ethanol Standards:**

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2015.02.10 11:00:04 -06:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

**Analyst:** \_\_\_\_\_

  
Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 2989.01**

AS IV Serial no: 087967  
Version no: 532B

TEST RECORD 00352

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/23/16 19:29 .000  
Calibration Check:  
19 02/23/16 19:29 .100

Subject Name

Subject I.D.

Operator Name, I.D.

*B PA 2004*

Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 532B

TEST RECORD 00353

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/23/16 19:31 .000  
Calibration Check:  
20 02/23/16 19:31 .099

Subject Name

Subject I.D.

Operator Name, I.D.

*B PA 2004*

Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 532B

TEST RECORD 00354

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
02/23/16 19:34 .000  
Calibration Check:  
21 02/23/16 19:34 .098

Subject Name

Subject I.D.

Operator Name, I.D.

*B PA 2004*

Location

*Joplin Jail*

AS IV Serial no: 087967  
Version no: 532B

TEST RECORD 00355

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 02/23/16 19:35

Subject Name

Subject I.D.

Operator Name, I.D.

*B PA 2004*

Location

*Joplin Jail*