



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 087974	PRINTER SN 08C.3556.372	DATE OF INSPECTION 05/20/2016
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 203 N 2nd Street, Bates City	TIME OF INSPECTION 12:59 pm
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 14220 EXP. DATE 09/24/2016

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2315 SIMULATOR EXP DATE 05/19/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .098

TEST 2 → .098

TEST 3 → .099

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE Virginia DeBoard	PRINT NAME Virginia DeBoard
TYPE II PERMIT NUMBER/EXPIRATION DATE 25004 / 02/10/2017	TELEPHONE NUMBER (816) 881-4617

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

f

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00055

Temp Date Time ^{2/} 210L

Air Blank:
05/20/16 12:59 .000
Calibration Check:
24 05/20/16 12:59 .000

Subject Name

Subject I.D.

Virginia DeBoard 986

Operator Name, I.D.

203 N 2nd St

Location

Bates City MO

f=

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00056

Temp Date Time ^{2/} 210L

Air Blank:
05/20/16 13:01 .000
Calibration Check:
24 05/20/16 13:01 .000

Subject Name

Subject I.D.

Virginia DeBoard 986

Operator Name, I.D.

203 N 2nd St

Location

Bates City MO

f

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00057

Temp Date Time ^{2/} 210L

Air Blank:
05/20/16 13:03 .000
Calibration Check:
25 05/20/16 13:03 .000

Subject Name

Subject I.D.

Operator Name, I.D.

Virginia DeBoard 986

Location

203 N 2nd St

Bates City MO

n

AS IV Serial no: 087974
Version no: 532B

TEST RECORD 00058

Temp Date Time ^{2/} 210L

VOID: RFI
12 05/20/16 13:03

Subject Name

Subject I.D.

Virginia DeBoard 986

Operator Name, I.D.

203 N 2nd St

Location

Bates City MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **14220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 25, 2014**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 24, 2016** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
VIRGINIA DEBOARD

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/10/2015

NUMBER 250041

EXPIRES 2/10/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator DEBOARD, VIRGINIA
Permit No 250041
Date Issued 2/10/2015 Date Expires 2/10/2017