



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Brian Lutmer at 9:11 am, Jan 15, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 095961	PRINTER SN 099.3586.643	DATE OF INSPECTION 01/14/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 715 W. Center Circle		TIME OF INSPECTION 10:24 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing LOT # 14001 EXP. DATE 04/30/2016
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2731 SIMULATOR EXP DATE 08/12/2016

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .101	TEST 2  .101	TEST 3  .101
--------------	--------------	--------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)    2	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	--------------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME E. Jason Hartsell
TYPE II PERMIT NUMBER/EXPIRATION DATE 260005 01/06/2018	TELEPHONE NUMBER (417) 725-2510

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

# Nixa Police Department

## Calibration Check slip's

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00917

Temp	Date	Time	s/
		210L	

Air Blank:  
01/14/16 10:25 .000  
Calibration Check:  
25 01/14/16 10:25 .101

Subject Name

Test 1

Subject I.D.

*G. J. [Signature]* 417  
Operator Name, I.D.

Location

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00918

Temp	Date	Time	s/
		210L	

Air Blank:  
01/14/16 10:27 .000  
Calibration Check:  
26 01/14/16 10:27 .101

Subject Name

Test 2

Subject I.D.

*G. J. [Signature]* 417  
Operator Name, I.D.

Location

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00919

Temp	Date	Time	s/
		210L	

Air Blank:  
01/14/16 10:29 .000  
Calibration Check:  
26 01/14/16 10:29 .101

Subject Name

Test 3

Subject I.D.

*G. J. [Signature]* 417  
Operator Name, I.D.

Location

# Nixa Police Department

## RFI Evidence slip

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00920

Temp Date Time 210L  
-----

VOID: RFI  
12 01/14/16 10:30

Subject Name

*RFI*

Subject I.D.

*E. J. [Signature] 487*

Operator Name, I.D.

Location

-----

-----

# Nixa Police Department

## Blank (Zero) test Evidence slip

AS IV Serial no: 095961  
Version no: 532B

TEST RECORD 00921

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:

01/14/16 10:32 .000

Subject Test: Auto

27 01/14/16 10:32 .000

Subject Name

*Blank Test*

Subject I.D.

*E. J. Anderson 417*

Operator Name, I.D.

Location

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 14001**

**EXPIRATION DATE: April 30, 2016 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014  
The expiration date for this lot number is April 30, 2016 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**ERIC J HARTSELL**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/6/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 260005

EXPIRES 1/6/2018

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator HARTSELL, ERIC  
 Permit No 260005  
 Date Issued 1/6/2016 Date Expires 1/6/2018