



RECEIVED

By Carol Day at 10:07 am, Feb 11, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 096228	PRINTER SN 097-3584-044	DATE OF INSPECTION 02/10/2016
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) NORMANDY PD, 7700 NATURAL BRIDGE, NORMANDY, MO	TIME OF INSPECTION 10:53 pm
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LOT # 15120 EXP. DATE 04/29/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD1615 SIMULATOR EXP DATE 07/06/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101

TEST 2 .102

TEST 3 .102

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

UNIT WAS SENT TO INTOXIMETERS FOR REPAIR. THE PIN SOCKET #52 WAS REPLACED.

A CALIBRATION WAS ALSO DONE

INSPECTING OFFICER

SIGNATURE 	PRINT NAME BRANDON W LANGE
---------------	-------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 250161 7/22/2017	TELEPHONE NUMBER (314) 385-3300
---	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 00413

Temp Date Time ^{s/} 210L

Air Blank:
02/10/16 22:53 .000
Subject Test: Auto
23 02/10/16 22:53 .000

Subject Name

Jange

Subject I.D.

Jange 206

Operator Name, I.D.

Jange 206

Location

Normandy PD

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 00415

Temp Date Time ^{s/} 210L

Air Blank:
02/10/16 22:58 .000
Calibration:
24 02/10/16 22:58 .100

Subject Name

Cal: Draft

Subject I.D.

Operator Name, I.D.

Jange 206

Location

Normandy PD

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 00416

Temp Date Time ^{s/} 210L

Air Blank:
02/10/16 23:00 .000
Calibration Check:
25 02/10/16 23:00 .101

Subject Name

Mant 1

Subject I.D.

N/A

Operator Name, I.D.

Jange 206

Location

Normandy PD

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 00417

Temp Date Time ^{s/} 210L

Air Blank:
02/10/16 23:02 .000
Calibration Check:
25 02/10/16 23:02 .102

Subject Name

Mant 2

Subject I.D.

N/A

Operator Name, I.D.

Jange 206

Location

Normandy PD

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 00418

Temp Date Time ^{s/} 210L

Air Blank:
02/10/16 23:04 .000
Calibration Check:
26 02/10/16 23:04 .102

Subject Name

Mant 3

Subject I.D.

N/A

Operator Name, I.D.

Jange 206

Location

Normandy PD

AS IV Serial no: 096228
Version no: 532B

TEST RECORD 00419

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/10/16 23:05

Subject Name

B.F.J

Subject I.D.

N/A

Operator Name, I.D.

Jange 206

Location

Normandy PD



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
 TYPE II


BRANDON W LANGE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015
 NUMBER 250161
 EXPIRES 7/22/2017


 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY
 Gal Vesterby
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
 L&S 4 (R8-10)

MO 593-0771 (9-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator LANGE, BRANDON
 Permit No 250161
 Date Issued 7/22/2015 Date Expires 7/22/2017



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.