

**RECEIVED**

By Carol Day at 1:33 pm, May 26, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097400	PRINTER SN 096.3580.880	DATE OF INSPECTION 05/25/2016
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Doniphan Police Department 124 Jefferson Street, Doniphan, Missouri 63935	TIME OF INSPECTION 10:45 am
---	--------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER RepCo Marketing Inc. LOT # 15001 EXP. DATE 05/20/2017 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP2209 SIMULATOR EXP DATE 05/17/2017 CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .102

TEST 2 ← .101

TEST 3 ← .101

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 | (0-.04) 40 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument operated within DHSS standards.

**INSPECTING OFFICER**

SIGNATURE

*S.K. Talburt*

PRINT NAME

Shayne K. Talburt

TYPE II PERMIT NUMBER/EXPIRATION DATE

250314 12/28/2017

TELEPHONE NUMBER

(573) 840-9500

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial not 652400  
Version not 502B

TEST RECORD 00130

Temp Date Time 2101

Air Driekt  
05/25/16 10:57 .000  
Calibration Check  
24 05/25/16 10:57 .101

Subject Name

MAINT TEST

Subject J.N.

Operator Name: J.N.

S.K. TALBURT # 1110  
Location

124 JEFFERSON ST

DONIPHAN, MO 63935

AS IV Serial not 652400  
Version not 502B

TEST RECORD 00130

Temp Date Time 2101

Air Driekt  
05/25/16 10:49 .000  
Calibration Check  
23 05/25/16 10:49 .102

Subject Name

MAINT TEST

Subject J.N.

Operator Name: J.N.

S.K. TALBURT #1110  
Location

124 JEFFERSON STREET,

DONIPHAN, MO 63935

AS IV Serial not 652400  
Version not 502B

TEST RECORD 00130

Temp Date Time 2101

Air Driekt  
05/25/16 10:53 .000  
Calibration Check  
24 05/25/16 10:53 .103

Subject Name

MAINT TEST

Subject J.N.

Operator Name: J.N.

S.K. TALBURT #1110  
Location

124 JEFFERSON STREET

DONIPHAN, MO 63935

AS IV Serial not 652400  
Version not 502B

TEST RECORD 00130

Temp Date Time 2300

WOTH RT  
12 05/25/16 13:00

Subject Name

MAINT TEST

Subject J.N.

Operator Name: J.N.

S.K. TALBURT #1110  
Location

124 JEFFERSON ST

DONIPHAN, MO 63935

**REPCO MARKETING INC.**3101-188 STONYBROOK DRIVE  
RALEIGH, N.C. 27604  
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at  
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT  
TYPE II**

**SHAYNE K TALBURT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250314

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



**Operator** TALBURT, SHAYNE  
**Permit No** 250314  
**Date Issued** 12/28/2015 **Date Expires** 12/28/2017