



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097400	PRINTER SN 096.3580.880	DATE OF INSPECTION 07/29/2016
LOCATION OF INSTRUMENT (STREET AND CITY) Doniphan Police Department 124 Jefferson Street, Doniphan, Missouri 63935		TIME OF INSPECTION 4:19 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER RepCo Marketing Inc. LOT # 15001 EXP. DATE 05/20/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIMULATOR SN MP2209 SIMULATOR EXP DATE 05/17/2017

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .102	TEST 2 ← .102	TEST 3 ← .102
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

This instrument operated within DHSS standards.

INSPECTING OFFICER

SIGNATURE <i>S.K. Talburt</i>	PRINT NAME Shayne K. Talburt
TYPE II PERMIT NUMBER/EXPIRATION DATE 250314 12/28/2017	TELEPHONE NUMBER (573) 840-9500

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

±

AS IV Serial no: 097400
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00142

Temp Date Time 210L s/

Air Blank:

07/29/16 16:25 .000

Calibration Check: 27 07/29/16 16:25 .102

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT #1110

Location

DOMIPHAN POLICE

DEPARTMENT

±

AS IV Serial no: 097400
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00143

Temp Date Time 210L s/

Air Blank:

07/29/16 16:28 .000

Calibration Check: 27 07/29/16 16:28 .102

Subject Name

S.K. TALBUNT

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT #1110

Location

DOMIPHAN POLICE

DEPARTMENT

±

AS IV Serial no: 097400
Version no: 532B

TEST RECORD 00144

Temp Date Time 210L s/

Air Blank:

07/29/16 16:31 .000

Calibration Check: 28 07/29/16 16:31 .102

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT #1110

Location

DOMIPHAN POLICE

DEPARTMENT

±

AS IV Serial no: 097400
Version no: 532B

TEST RECORD 00145

Temp Date Time 210L s/

VOID: RFI

12 07/29/16 16:32

Subject Name

MAINT TEST

Subject I.D.

Operator Name, I.D.

S.K. TALBUNT #1110

Location

DOMIPHAN POLICE

DEPARTMENT

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.