RECEIVED
By Carol Day at 11:20 am, Jul 13, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
097405	08C. 3556, 388	7-5-16
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
Steele Police Department 115 S. Walnut Steele, MO.		0513

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [x] SIMULATOR SOLUTION
- [ ] COMPRESSED ETHANOL-GAS MIXTURE
- [x] STANDARD SUPPLIER Guth LAB. LOT # 16080 EXP. DATE 3-7-18
- [x] SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD1758 SIMULATOR EXP DATE 10-20-16
- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1	TEST 2	TEST 3
.102	.101	.102

- [x] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
		1		1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

machine working to specification

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Josh Carter	JOSH CARTER
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240348 10-3-16	573-695-2100

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

JUL. 5. 2016 5:37AM STEELE P.D. No. 0652 P. 3

AS IV Serial no: 097405
Version no: 5328

TEST RECORD 00315

Temp	Date	Time	210L

Air Blank:
07/05/16 05:23 .000
Calibration Check:
24 07/05/16 05:23 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097405
Version no: 5328

TEST RECORD 00317

Temp	Date	Time	210L

Air Blank:
07/05/16 05:26 .000
Calibration Check:
24 07/05/16 05:26 .102

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097405
Version no: 5328

TEST RECORD 00318

Temp	Date	Time	210L

NOTE: NFI
24 07/05/16 05:27

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097405
Version no: 5328

TEST RECORD 00316

Temp	Date	Time	210L

Air Blank:
07/05/16 05:25 .000
Calibration Check:
24 07/05/16 05:25 .101

Subject Name

Subject I.D.

Operator Name, I.D.

Location

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **16080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 9, 2016,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1210%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 7, 2018** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN08051301** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

JOSHUA S CARTER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following Breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014

NUMBER 240348

EXPIRES 10/3/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald Vasterling, acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CARTER, JOSHUA
Permit No 240348
Date Issued 10/3/2014 Date Expires 10/3/2016