RECEIVED
By Brian Lutmer at 4:57 pm, Jan 15, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
097408	096.3580.872	01/04/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 103 ST, HWY H, MINER,MO		TIME OF INSPECTION 10:36 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- [x] DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- [x] TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- [x] PRINTER WORKING PROPERLY
- [x] TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- [x] SIMULATOR SOLUTION
- [ ] COMPRESSED ETHANOL-GAS MIXTURE

- [x] STANDARD SUPPLIER REPCO MARKETING LOT # 14001 EXP. DATE 04/30/2016

- [x] SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2273 SIMULATOR EXP DATE 12/11/2016

- [x] CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
  Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - [x] 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - [ ] 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - [ ] 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .101	TEST 2 ☛ .101	TEST 3 ☛ .100

- [x] RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

OPERATING WITHIN DOHSS STANDARDS

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸	LANCE ASH
TYPE II PERMIT NUMBER/EXPIRATION DATE 250319 12-28-2017	TELEPHONE NUMBER (573) 471-8568

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number **14001** of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain **.1216** gms/dl +/-.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interferring substance.

This solution will produce a vapor alcohol value of **.100** +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is **May 1, 2014** The expiration date for this lot number is **April 30, 2016** at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.

Form RM 02

AS IV Serial no: 097408
Version no: 532B

TEST RECORD 00104

Temp	Date	Time	g/210L

Air Blank:
01/04/16 22:40 .000
Calibration Check:
24 01/04/16 22:40 .101

Subject Name
Test 2
Subject I.D.
ASL 250319
Operator Name, I.D.

Location

AS IV Serial no: 097408
Version no: 532B

TEST RECORD 00103

Temp	Date	Time	g/210L

Air Blank:
01/04/16 22:38 .000
Calibration Check:
24 01/04/16 22:38 .101

Subject Name
Test 1
Subject I.D.
ASL 250319
Operator Name, I.D.

Location

AS IV Serial no: 097408
Version no: 532B

TEST RECORD 00102

Temp	Date	Time	g/210L

Air Blank:
01/04/16 22:36 .000
Subject Test: Auto
23 01/04/16 22:36 .000

Subject Name
Sober Sample
Subject I.D.
ASL 250319
Operator Name, I.D.

Location

AS IV Serial no: 097408
Version no: 532B

TEST RECORD 00106

Temp	Date	Time	g/210L

VOID: RFI
12 01/04/16 22:43

Subject Name
RFI
Subject I.D.
ASL 250319
Operator Name, I.D.

Location

AS IV Serial no: 097408
Version no: 532B

TEST RECORD 00105

Temp	Date	Time	g/210L

Air Blank:
01/04/16 22:42 .000
Calibration Check:
25 01/04/16 22:42 .100

Subject Name
Test 3
Subject I.D.
ASL 250319
Operator Name, I.D.

Location

2

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT
TYPE II

LANCE ASH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250319

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R5-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ASH, LANCE
Permit No 250319
Date Issued 12/28/2015 Date Expires 12/28/2017