

REPCO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-578-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 14001

EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President.
RepCo Marketing, Inc.

AS IU Serial no: 097408
Version no: 532B

TEST RECORD 00166

Temp Date Time 210L

Air Blank:
03/15/16 22:12 .000
Calibration Check:
21 03/15/16 22:12 .099

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

Ash 250319

Location

Miner PD

AS IU Serial no: 097409
Version no: 532B

TEST RECORD 00170

Temp Date Time 210L

VOID: RFI
12 03/15/16 22:18

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Ash 250319

Location

Miner PD

AS IU Serial no: 097408
Version no: 532B

TEST RECORD 00169

Temp Date Time 210L

Air Blank:
03/15/16 22:17 .000
Subject Test: Auto
24 03/15/16 22:17 .000

Subject Name

Sober Sample

Subject I.D.

Operator Name, I.D.

Ash 250319

Location

Miner PD

AS IU Serial no: 097408
Version no: 532B

TEST RECORD 00168

Temp Date Time 210L

Air Blank:
03/15/16 22:15 .000
Calibration Check:
24 03/15/16 22:15 .097

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

Ash 250319

Location

Miner PD

AS IU Serial no: 097408
Version no: 532B

TEST RECORD 00167

Temp Date Time 210L

Air Blank:
03/15/16 22:14 .000
Calibration Check:
23 03/15/16 22:14 .097

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

Ash 250319

Location

Miner PD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
LANCE ASH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250319

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (8-10)

LAB-4 (88-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ASH, LANCE
Permit No 250319
Date Issued 12/28/2015 Date Expires 12/28/2017