

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 15001
EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 897408
Version no: 532B

TEST RECORD 00189

Temp Date Time ^{s/} 210L

Air Blank:
06/25/16 10:21 .000
Calibration Check:
23 06/25/16 10:21 .101

Subject Name

Test #2
Subject I.D.

Operator Name, I.D.

AsL 250319
Location
Miner PD

AS IV Serial no: 897408
Version no: 532B

TEST RECORD 00188

Temp Date Time ^{s/} 210L

Air Blank:
06/25/16 10:19 .000
Calibration Check:
22 06/25/16 10:19 .102

Subject Name

Test 1
Subject I.D.

Operator Name, I.D.

AsL 250319
Location
Miner PD

AS IV Serial no: 897408
Version no: 532B

TEST RECORD 00187

Temp Date Time ^{s/} 210L

Air Blank:
06/25/16 10:17 .000
Subject Test: Auto
22 06/25/16 10:17 .000

Subject Name

Sober Sample
Subject I.D.

Operator Name, I.D.

AsL
Location
Miner PD

AS IV Serial no: 897408
Version no: 532B

TEST RECORD 00191

Temp Date Time ^{s/} 210L

Air Blank:
06/25/16 10:20

Subject Name

RFI Test
Subject I.D.

Operator Name, I.D.

AsL 250319
Location
Miner PD

AS IV Serial no: 897408
Version no: 532B

TEST RECORD 00190

Temp Date Time ^{s/} 210L

Air Blank:
06/25/16 10:23 .000
Calibration Check:
24 06/25/16 10:23 .101

Subject Name

Test 3
Subject I.D.

Operator Name, I.D.

AsL 250319
Location
Miner PD



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
LANCE ASH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 806.111 through 806.119 RSMo.

DATE 12/28/2015

NUMBER 250319

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4.(06-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **ASH, LANCE**
 Permit No **250319**
 Date Issued **12/28/2015** Date Expires **12/28/2017**