RECEIVED
By Brian Lutmer at 4:59 pm, Jan 15, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
097413	091.3580.067	01/04/2016
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
52 YOUNG DR CALVERTON PARK MO 63135		14.05 hrs

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER GUTH LABORATORIES LOT # 15220 EXP. DATE 09/28/2017

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD3327 SIMULATOR EXP DATE 01/21/2016

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .099	TEST 2 ☛ .099	TEST 3 ☛ .098

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	2	(.15-.19)	0	(OVER .19)	2

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Guth lab lot num 15220 exp 09/28/2017

INSPECTING OFFICER

SIGNATURE	PRINT NAME
Lowell Moore	Lowell Moore
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
240234 05/12/2016	(314) 524-1212

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 30, 2015,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 28, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
*Testing was conducted using Cerilliant Reference Standard lot number **FN08051301** whose values are traceable to NIST.*
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

3

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00605

Temp	Date	Time	s/ 210L
Air Blank:			
	01/04/16	14:28	.000
Calibration Check:			
22	01/04/16	14:28	.098

Subject Name

Subject I.D.

L Moore 172

Operator Name, I.D.

Location

2

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00604

Temp	Date	Time	s/ 210L
Air Blank:			
	01/04/16	14:27	.000
Calibration Check:			
22	01/04/16	14:27	.099

Subject Name

Subject I.D.

L Moore 172

Operator Name, I.D.

Location

1

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00603

Temp	Date	Time	s/ 210L
Air Blank:			
	01/04/16	14:25	.000
Calibration Check:			
21	01/04/16	14:25	.099

Subject Name

Subject I.D.

L. Moore 239240

Operator Name, I.D.

Location

RFI

AS IV Serial no: 097413
Version no: 004C

TEST RECORD 00606

Temp	Date	Time	s/ 210L
Void: RFI			
12	01/04/16	14:29	

Subject Name

Subject I.D.

L. Moore 173

Operator Name, I.D.

Location

Copy

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

LOWELL J MOORE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/12/2014

NUMBER 240234

EXPIRES 5/12/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (8-10)

LAB-4 (R8-10)

COPY

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOORE, LOWELL
Permit No 240234
Date Issued 5/12/2014 Date Expires 5/12/2016