

AS IV Serial no: 097415
Version no: 532B

TEST RECORD 00027^{g/}
Temp Date Time 210L

Air Start
04/11/16 20:58 .000
Calibration Check
04/11/16 20:58 .000

Printed: g/

04/11/16

Operator Name: I.D.

T.A. Hadlock 789

ZONE 21 OFFICE

AS IV Serial no: 097415
Version no: 532B

TEST RECORD 00028^{g/}
Temp Date Time 210L

Air Start
04/11/16 20:59 .000
Calibration Check
04/11/16 20:59 .000

Printed: g/

04/11/16

Operator Name: I.D.

T.A. Hadlock 789

ZONE 21 OFFICE

AS IV Serial no: 097415
Version no: 532B

TEST RECORD 00029^{g/}
Temp Date Time 210L

Air Start
04/11/16 21:01 .000
Calibration Check
04/11/16 21:01 .000

Printed: g/

04/11/16

Operator Name: I.D.

T.A. Hadlock 789

ZONE 21 OFFICE

AS IV Serial no: 097415
Version no: 532B

TEST RECORD 00030^{g/}
Temp Date Time 210L

Air Start
04/11/16 21:02

Printed: g/

04/11/16

Operator Name: I.D.

T.A. Hadlock 789

ZONE 21

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 14001
EXPIRATION DATE: April 30, 2016 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 14001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1216 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.


This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 1, 2014
The expiration date for this lot number is April 30, 2016 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
TODD A HADLOCK



is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repair and operate the following breath analyzer(s):

DATAMASTER, INTOX DMT, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of section 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/7/2014
NUMBER 240054
EXPIRES 3/7/2016

MO 58007 PL (6-10)


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (05-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **HADLOCK, TODD**
Permit No **240054**
Date Issued **3/7/2014** Date Expires **3/7/2016**