

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00201

Temp Date Time ^{s/} 210L

Air Blank:
02/01/16 11:32 .000
Calibration Check:
19 02/01/16 11:32 .103

Subject Name

CALIBRATION CHECK

Subject I.D.

Operator Name, I.D.

Ronald Chryf *501

Location

200 N MAIN ST

PALMYRA MO

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00202

Temp Date Time ^{s/} 210L

Air Blank:
02/01/16 11:34 .000
Calibration Check:
19 02/01/16 11:34 .103

Subject Name

CALIBRATION CHECK

Subject I.D.

Operator Name, I.D.

Ronald Chryf 501

Location

200 N MAIN ST

PALMYRA MO

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00203

Temp Date Time ^{s/} 210L

Air Blank:
02/01/16 11:35 .000
Calibration Check:
20 02/01/16 11:35 .104

Subject Name

CALIBRATION CHECK

Subject I.D.

Operator Name, I.D.

Ronald Chryf 501

Location

200 N MAIN

PALMYRA MO

AS IV Serial no: 097419
Version no: 532B

TEST RECORD 00204

Temp Date Time ^{s/} 210L

VOID: RFI
12 02/01/16 11:36

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

Ronald Chryf 501

Location

200 N MAIN

PALMYRA MO



GUTH LABORATORIES, INC.

598 NORTH 62ND STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-664-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator.

Random Samples of Lot Number 15050 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on March 11, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1218% (w/vol) ethyl alcohol. The expiration date for this lot number is March 9, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- 0.2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number EN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

RONALD C PEER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER


for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119, RSMo.

DATE 7/22/2014

NUMBER 240306

EXPIRES 7/22/2016

MO 500-2771 (2-13)


DIRECTOR OF STATE PUBLIC HEALTH ASSISTANT
Gail Vaskerly
acting director
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (2-13)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator **PEER, RONALD**
Permit No **240306**
Date Issued **7/22/2014** Date Expires **7/22/2016**