RECEIVED
By Carol Day at 4:04 pm, Feb 16, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
097421	096.3580.987	02/07/2016

LOCATION OF INSTRUMENT (STREET AND CITY)	TIME OF INSPECTION
110 Silverlady Lane Branson West Mo,65737	3:10 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth Laboratories LOT # 15120 EXP. DATE 04/29/2017

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD3139 SIMULATOR EXP DATE 07/24/2016

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .099	TEST 2 ☛ .101	TEST 3 ☛ .101

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	1	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE	PRINT NAME
▸ Vincent Collins	Vincent Collins
TYPE II PERMIT NUMBER/EXPIRATION DATE	**TELEPHONE NUMBER**
240349 10-03-2016	(417) 272-3400

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

GUTH LABORATORIES, INC.
690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT

TYPE II

VINCENT COLLINS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 10/3/2014

NUMBER 240349

EXPIRES 10/3/2016

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Gail Vasterling, acting director

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (06-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator COLLINS, VINCENT
Permit No 240349
Date Issued 10/3/2014 Date Expires 10/3/2016

Test #1

AS IV Serial no: 097421
Version no: 532B

TEST RECORD 00255

Temp	Date	Time	g/210L
Air Blank:			
	02/07/16	15:10	.000
Calibration Check:			
20	02/07/16	15:10	.099

Subject Name

Test #1

Subject I.D.

Vincent Collins 240349

Operator Name, I.D.

Branson West P.D

Location

110 Silverlady Ln

Branson West MO 65737

Vincent Collins

Test #2

AS IV Serial no: 097421
Version no: 532B

TEST RECORD 00256

Temp	Date	Time	g/210L
Air Blank:			
	02/07/16	15:12	.000
Calibration Check:			
20	02/07/16	15:12	.101

Subject Name

Test #2

Subject I.D.

Vincent Collins 240349

Operator Name, I.D.

Branson West P.D.

Location

110 Silverlady Ln

Branson West MO 65737

Vincent Collins

Test #3

AS IV Serial no: 097421
Version no: 532B

TEST RECORD 00257

Temp	Date	Time	g/210L
Air Blank:			
	02/07/16	15:13	.000
Calibration Check:			
21	02/07/16	15:13	.101

Subject Name

Test #3

Subject I.D.

Vincent Collins 240349

Operator Name, I.D.

Branson West P.D.

Location

110 Silverlady Ln

Branson West MO 65737

Vincent Collins

Test #4

AS IV Serial no: 097421
Version no: 532B

TEST RECORD 00258

Temp	Date	Time	g/210L
VOID: RFI			
12	02/07/16	15:15	

Subject Name

RFI Test

Subject I.D.

Vincent Collins 240349

Operator Name, I.D.

Branson West P.D.

Location

110 Silverlady Ln

Branson West MO 65737

Vincent Collins

Test #5

AS IV Serial no: 097421
Version no: 532B

TEST RECORD 00259

Temp	Date	Time	g/210L
Air Blank:			
	02/07/16	15:16	.000
Subject Test: Auto			
22	02/07/16	15:16	.000

Subject Name

Blank Test

Subject I.D.

Vincent Collins 240349

Operator Name, I.D.

Branson West PD

Location

110 Silverlady Ln

Branson West MO 65737

Vincent Collins