



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:29 am, Jul 05, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <i>097424</i>	PRINTER SN <i>096.3580.984</i>	DATE OF INSPECTION <i>07/04/2016</i>
LOCATION OF INSTRUMENT (STREET AND CITY) <i>Wentzville Police Dept. 1019 Schroeder Creek Blvd - Wentzville, MO 63385</i>		TIME OF INSPECTION <i>0707</i>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER *Intoximeters* LOT # *AG533902* EXP. DATE *12/05/2017*
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

- CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <i>.097</i>	TEST 2 • <i>.096</i>	TEST 3 • <i>.096</i>
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <i>3</i>	(0-.04) <i>∅</i>	(.05-.09) <i>∅</i>	(.10-.14) <i>∅</i>	(.15-.19) <i>∅</i>	(OVER .19) <i>∅</i>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

True-Cal II : 0.098

INSPECTING OFFICER	
SIGNATURE <i>K. Bruns</i>	PRINT NAME <i>Kyle Bruns</i>
TYPE II PERMIT NUMBER/EXPIRATION DATE <i>250285 / 11-23-2017</i>	TELEPHONE NUMBER <i>636-327-5105</i>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097424
Version no: 532B

TEST RECORD 00521

Temp Date Time ^{s/} 210L

Air Blank:
07/04/16 07:07 .000
Calibration Check:
23 07/04/16 07:07 .097

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097424
Version no: 532B

TEST RECORD 00522

Temp Date Time ^{s/} 210L

Air Blank:
07/04/16 07:09 .000
Calibration Check:
24 07/04/16 07:09 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097424
Version no: 532B

TEST RECORD 00523

Temp Date Time ^{s/} 210L

Air Blank:
07/04/16 07:11 .000
Calibration Check:
25 07/04/16 07:11 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 097424
Version no: 532B

TEST RECORD 00524

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/04/16 07:12

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 7-Dec-2015

Lot # AG533902 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
5-Dec-2017	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010595	208.9 ppm
EB0010561	103.7 ppm	EB0010562	104.9 ppm
EB0010681	52.22 ppm	EB0010579	52.94 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2015.12.07 14:43:27 -06:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Analyst: 
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KYLE T BRUNS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/23/2015

NUMBER 250285

EXPIRES 11/23/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator BRUNS, KYLE
 Permit No 250285
 Date Issued 11/23/2015 Date Expires 11/23/2017