RECEIVED
By Ellen Strawsine at 2:34 pm, Jan 20, 2016

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN	PRINTER SN	DATE OF INSPECTION
097425	096.3580.866	01/17/2016
LOCATION OF INSTRUMENT (STREET AND CITY)		TIME OF INSPECTION
201 S. Chestnut st. Richland Mo. 65556		3:37 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- ☑ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- ☑ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- ☑ PRINTER WORKING PROPERLY
- ☑ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☑ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☑ STANDARD SUPPLIER Guth labortories LOT # 15220 EXP. DATE 09/28/2017

☑ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN DR6931 SIMULATOR EXP DATE 12/11/2016

☑ CALIBRATION CHECK – **(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- ☑ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- ☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- ☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ☛ .105	TEST 2 ☛ .104	TEST 3 ☛ .105

☑ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Unit operating within established limits

INSPECTING OFFICER

SIGNATURE	PRINT NAME
David Moser	David Moser
TYPE II PERMIT NUMBER/EXPIRATION DATE	TELEPHONE NUMBER
250316 / 12/28/2017	(573) 765-4144

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AN EQUAL OPPORTUNITY/AFFIRMATIVE ACTION EMPLOYER
services provided on a nondiscriminatory basis

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

DAVID L MOSER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/28/2015

NUMBER 250316

EXPIRES 12/28/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSER, DAVID
Permit No 250316
Date Issued 12/28/2015 Date Expires 12/28/2017

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00117

Temp	Date	Time	s/ 210L

Air Blank:
01/17/16 15:41 .000
Calibration Check:
26 01/17/16 15:41 .104

Subject Name
TEST

Subject I.D.
3

Operator Name, I.D.
DAVID MOSER
406/5003

Location
RICHLAND PD

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00116

Temp	Date	Time	s/ 210L

Air Blank:
01/17/16 15:39 .000
Calibration Check:
26 01/17/16 15:39 .104

Subject Name
TEST

Subject I.D.
2

Operator Name, I.D.
DAVID MOSER
406/5003

Location
RICHLAND PD

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00115

Temp	Date	Time	s/ 210L

Air Blank:
01/17/16 15:37 .000
Calibration Check:
25 01/17/16 15:37 .105

Subject Name
TEST 1

Subject I.D.
1

Operator Name, I.D.
DAVID MOSER
406/5003

Location
RICHLAND PD

AS IV Serial no: 097425
Version no: 532B

TEST RECORD 00118

Temp	Date	Time	s/ 210L

VOID: RFI
12 01/17/16 15:42

Subject Name
TEST RFI

Subject I.D.
4

Operator Name, I.D.
DAVID MOSER
406/5003

Location
RICHLAND PD

GUTH LABORATORIES, INC.
590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15220** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 30, 2015,** using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 28, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%.**

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.