

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00070

Temp	Date	Time	g/210L
Air Blank:	02/16/16	14:39	.000
Calibration:	27 02/16/16	14:39	.100

Subject Name CALIBRATION
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00071

Temp	Date	Time	g/210L
Air Blank:	02/16/16	14:57	.000
Calibration Check:	25 02/16/16	14:57	.100

Subject Name
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00072

Temp	Date	Time	g/210L
Air Blank:	02/16/16	14:59	.000
Calibration Check:	25 02/16/16	14:59	.100

Subject Name
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00073

Temp	Date	Time	g/210L
Air Blank:	02/16/16	15:02	.000
Calibration Check:	26 02/16/16	15:02	.099

Subject Name
MAINT TEST
Subject I.D.

Operator Name, I.D.
Jimmy Cleveland
Location
2920 N. Shamrock
Jefferson City

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00075

Temp	Date	Time	g/210L
VOID: RFI	12 02/16/16	15:05	

Subject Name
RFI TEST
Subject I.D.

Operator Name, I.D.
Jimmy L. Cleveland
Location
2920 N. Shamrock
Jefferson City



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

JIMMY L CLEVELAND

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/22/2015

NUMBER 250155

EXPIRES 7/22/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0/71 (6 10)

LAB 4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator CLEVELAND, JIMMY
Permit No 250155
Date Issued 7/22/2015 Date Expires 7/22/2017