

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00107

Temp Date Time ^{a/} 210L

Air Blank:
06/24/16 10:35 .000
Calibration Check:
23 06/24/16 10:35 .100

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

J.M. Huet 1374

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00109

Temp Date Time ^{a/} 210L

Air Blank:
06/24/16 10:39 .000
Calibration Check:
24 06/24/16 10:39 .101

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

J.M. Huet 1374

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00108

Temp Date Time ^{a/} 210L

Air Blank:
06/24/16 10:37 .000
Calibration Check:
23 06/24/16 10:37 .100

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

J.M. Huet

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 097427
Version no: 532B

TEST RECORD 00110 ^{a/}

Temp Date Time 210L

VOID: RFI
12 06/24/16 10:40

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

J.M. Huet 1374

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

REP CO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JASON M HURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/18/2016

NUMBER 260075

EXPIRES 2/18/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HURT, JASON
 Permit No 260075
 Date Issued 2/18/2016 Date Expires 2/18/2018



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097423	PRINTER SN 096.3580.985	DATE OF INSPECTION 06/24/2016
LOCATION OF INSTRUMENT (STREET AND CITY) Zone 10/11 office, Davis St. St. Clair		TIME OF INSPECTION 11:07 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing LOT # 15001 EXP. DATE 05/20/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP2128 SIMULATOR EXP DATE 03/03/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.102

TEST 2 ← 0.100

TEST 3 ← 0.102

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Jason M. Hurt
---------------	-----------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 260075/02/18/2018	TELEPHONE NUMBER (636) 300-2800
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

REP CO MARKETING INC.

3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.
LOT NUMBER: 15001
EXPIRATION DATE: May 20, 2017 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00033

Temp Date Time ^{g/} 210L

Air Blank:
06/24/16 11:07 .000
Calibration Check:
27 06/24/16 11:07 .102

Subject Name

TEST 1
Subject I.D.

Operator Name, I.D.

J.M. HURT 1374
Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00034

Temp Date Time ^{g/} 210L

Air Blank:
06/24/16 11:08 .000
Calibration Check:
27 06/24/16 11:08 .100

Subject Name

TEST 2
Subject I.D.

Operator Name, I.D.

J.M. HURT 1374
Location

ZONE 10/11 OFFICE

DAVIS ST ST. CLAIR

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00035

Temp Date Time ^{g/} 210L

Air Blank:
06/24/16 11:10 .000
Calibration Check:
28 06/24/16 11:10 .102

Subject Name

TEST 3
Subject I.D.

Operator Name, I.D.

J.M. HURT 1374
Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 097423
Version no: 532B

TEST RECORD 00036

Temp Date Time ^{g/} 210L

VOID: RFI
12 06/24/16 11:11

Subject Name

RFT
Subject I.D.

Operator Name, I.D.

J.M. HURT 1374
Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JASON M HURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/18/2016

NUMBER 260075

EXPIRES 2/18/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HURT, JASON
 Permit No 260075
 Date Issued 2/18/2016 Date Expires 2/18/2018



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102471	PRINTER SN 8A.3539.008	DATE OF INSPECTION 06/24/2016
LOCATION OF INSTRUMENT (STREET AND CITY) Zone 10/11 office, Davis St. St. Clair		TIME OF INSPECTION 11:26 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER RepCo Marketing LOT # 15001 EXP. DATE 05/20/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIMULATOR SN MP2128 SIMULATOR EXP DATE 03/03/2017

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.102

TEST 2 ← 0.104

TEST 3 ← 0.100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Jason M. Hurt
---------------	-----------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 260075/02/18/2018	TELEPHONE NUMBER (636) 300-2800
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 102471
Version no: 532B

TEST RECORD 00239

Temp Date Time a/
210L

Air Blank:
06/24/16 11:26 .000
Calibration Check:
26 06/24/16 11:26 .101

Subject Name

TEST 1

Subject I.D.

Operator Name, I.D.

J.M. HURT 1374

AS IV Serial no: 102471
Version no: 532B

TEST RECORD 00240

Temp Date Time a/
210L

Air Blank:
06/24/16 11:28 .000
Calibration Check:
26 06/24/16 11:28 .104

Subject Name

TEST 2

Subject I.D.

Operator Name, I.D.

J.M. HURT 1374

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 102471
Version no: 532B

TEST RECORD 00241

Temp Date Time a/
210L

Air Blank:
06/24/16 11:30 .000
Calibration Check:
26 06/24/16 11:30 .100

Subject Name

TEST 3

Subject I.D.

Operator Name, I.D.

J.M. HURT 1374

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

AS IV Serial no: 102471
Version no: 532B

TEST RECORD 00242

Temp Date Time a/
210L

Air Blank:
06/24/16 11:31

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

J.M. HURT 1374

Location

ZONE 10/11 OFFICE

DAVIS ST. ST. CLAIR

REPCO MARKETING INC.3101-188 STONYBROOK DRIVE
RALEIGH, N.C. 27604
919-876-5480**CERTIFICATE OF ANALYSIS****MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.****LOT NUMBER: 15001****EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015
The expiration date for this lot number is May 20, 2017 at
11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner, President
RepCo Marketing, Inc.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
JASON M HURT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/18/2016

NUMBER 260075

EXPIRES 2/18/2018

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator HURT, JASON
 Permit No 260075
 Date Issued 2/18/2016 Date Expires 2/18/2018