



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Carol Day at 3:23 pm, Feb 19, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097429	PRINTER SN 03A.2436.100	DATE OF INSPECTION 02/12/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 400 First Street Hillsboro, Mo. 63050		TIME OF INSPECTION 9:00 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Lab LOT # 15120 EXP. DATE 04/29/2017
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2222 SIMULATOR EXP DATE 03/05/2016

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .104	TEST 2 .104	TEST 3 .104
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	3	(0-.04)	(.05-.09)	1	(.10-.14)	1	(.15-.19)	(OVER .19)
----------	---	---------	-----------	---	-----------	---	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument Operates within the guidelines set forth by the Dept. of Health and Senior Services

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Cpl. Brian Taylor #365
TYPE/PERMIT NUMBER/EXPIRATION DATE 240139/04-03-2016	TELEPHONE NUMBER (636) 797-5000

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRIAN R TAYLOR

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

DATAMASTER, INTOX EC/IR II, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/3/2014

NUMBER 240139

EXPIRES 4/3/2016

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator TAYLOR, BRIAN
 Permit No 240139
 Date Issued 4/3/2014 Date Expires 4/3/2016

AS IV Serial no: 097429
Version no: 532B

TEST RECORD 00181

Temp Date Time 210L

Air Blank: 02/12/16 09:10 .000

Calibration Check: 24 02/12/16 09:10 .104

Subject Name

Text 3

Subject I.D.

Operator Name: I.D.

Cpl Brian Taylor 365

Location

400 First St

Hillsboro, MO 63050

AS IV Serial no: 097429
Version no: 532B

TEST RECORD 00180

Temp Date Time 210L

Air Blank: 02/12/16 09:08 .000

Calibration Check: 23 02/12/16 09:08 .104

Subject Name

Text 2

Subject I.D.

Operator Name: I.D.

Cpl Brian Taylor 365

Location

400 First St

Hillsboro, MO 63050

AS IV Serial no: 097429
Version no: 532B

TEST RECORD 00179

Temp Date Time 210L

Air Blank: 02/12/16 09:06 .000

Calibration Check: 23 02/12/16 09:06 .104

Subject Name

Text 1

Subject I.D.

Operator Name: I.D.

Cpl Brian Taylor 365

Location

400 First St

Hillsboro, MO 63050

AS IV Serial no: 097429
Version no: 532B

TEST RECORD 00178

Temp Date Time 210L

VOID: RTI 12 02/12/16 08:45

Subject Name

RFI Text

Subject I.D.

Operator Name: I.D.

Cpl Brian Taylor 365

Location

400 First St

Hillsboro, MO 63050



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 15120 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on May 4, 2015, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is April 29, 2017 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.