



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|--|----------------------------|----------------------------------|
| ALCO SENSOR IV SN 097438 | PRINTER SN 096.3580.921 | DATE OF INSPECTION 01/18/2016 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 9701 Marion Park Drive, Kanas City, MO 64137 | | TIME OF INSPECTION 1:41 am |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG521003 EXP. DATE 07/29/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 ● .101 | TEST 2 ● .102 | TEST 3 ● .102 |
|---------------|---------------|---------------|

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

| | |
|--|------------------------------------|
| SIGNATURE <i>Kori Smeiska</i> | PRINT NAME Kori Smeiska |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 250133/6-8-2017 | TELEPHONE NUMBER (816) 234-5000 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II

KORI SMEISKA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.


DATE 6/8/2015

NUMBER 250133

EXPIRES 6/8/2017

MO 550-0771 (6-10)


DIRECTOR OF STATE PUBLIC HEALTH LABORATORY


DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (16-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath from of expired air in Missouri.



Operator: SMEISKA, KORI
Permit No: 250133
Date Issued 6/8/2015 Date Expires 6/8/2017



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 3-Aug-2015

Lot # AG521003 Model 108cadd


| <u>Exp. Date</u> | <u>Cyl. Type</u> | <u>Component</u> | <u>Certified Concentration</u> |
|------------------|------------------|------------------|--------------------------------|
| 29-Jul-2017 | 108 | Ethanol | 0.100 ± 2% BrAC (272 ppm) |
| | | Nitrogen | Balance |

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

| <u>Serial No.</u> | <u>Concentration</u> | <u>Serial No.</u> | <u>Concentration</u> |
|-------------------|----------------------|-------------------|----------------------|
| EB0010581 | 391.8 ppm | EB0010603 | 392.5 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.9 ppm |
| EB0010285 | 209.0 ppm | EB0010595 | 208.9 ppm |
| EB0010561 | 103.7 ppm | EB0010562 | 104.9 ppm |
| EB0010681 | 52.22 ppm | EB0010579 | 52.94 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2015.08.04 09:52:31 -0500
Reason: Dry gas standard certification of analysts
Location: Airgas USA LLC (Lab)

Analyst: 
Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00160

Temp Date Time ^{s/} 210L

Air Blank:
01/18/16 01:41 .000
Calibration Check:
24 01/18/16 01:41 .101

Subject Name

Subject I.D.

Operator Name, I.D.

P.O. Smeiska #5260

Location

DUI Section

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00161

Temp Date Time ^{s/} 210L

Air Blank:
01/18/16 01:43 .000
Calibration Check:
25 01/18/16 01:43 .102

Subject Name

Subject I.D.

Operator Name, I.D.

P.O. Smeiska #5260

Location

DUI Section

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00162

Temp Date Time ^{s/} 210L

Air Blank:
01/18/16 01:45 .000
Calibration Check:
26 01/18/16 01:45 .102

Subject Name

Subject I.D.

Operator Name, I.D.

P.O. Smeiska #5260

Location

DUI Section

AS IV Serial no: 097438
Version no: 532B

TEST RECORD 00163

Temp Date Time ^{s/} 210L

VOID: RFI
12 01/18/16 01:47

Subject Name

Subject I.D.

Operator Name, I.D.

P.O. Smeiska #5260

Location

DUI Section