



AS IV Serial no: 097445  
Version no: 532B

TEST RECORD 00280

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/03/16 13:35 .000  
Calibration Check:  
22 06/03/16 13:35 .101

Subject Name

Subject I.D.

Operator Name, I.D.

DEUTMAN 885

Location

CCPD BATVAN

AS IV Serial no: 097445  
Version no: 532B

TEST RECORD 00281

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/03/16 13:37 .000  
Calibration Check:  
22 06/03/16 13:37 .101

Subject Name

Subject I.D.

Operator Name, I.D.

DEUTMAN 885

Location

CCPD BATVAN

AS IV Serial no: 097445  
Version no: 532B

TEST RECORD 00282

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
06/03/16 13:39 .000  
Calibration Check:  
23 06/03/16 13:39 .100

Subject Name

Subject I.D.

Operator Name, I.D.

DEUTMAN 885

Location

CCPD BATVAN

AS IV Serial no: 097445  
Version no: 532B

TEST RECORD 00283

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 06/03/16 13:40

Subject Name

Subject I.D.

Operator Name, I.D.

DEUTMAN 885

Location

CCPD BATVAN



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

**PERMIT  
 TYPE II**

**MICHAEL T DEUTMAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/6/2016

NUMBER 260007

EXPIRES 1/6/2018

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator DEUTMAN, MICHAEL  
 Permit No 260007  
 Date issued 1/6/2016 Date Expires 1/6/2018

**CERTIFICATE OF ANALYSIS**

**MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.**

**LOT NUMBER: 15001**

**EXPIRATION DATE: May 20, 2017 at 11:59 p.m.**

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 15001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1206 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L Breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is May 21, 2015  
The expiration date for this lot number is May 20, 2017 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marketing, Inc.