



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE

RECEIVED
By Carol Day at 7:54 am, Apr 20, 2016

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventive maintenance. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 097460	PRINTER SN 096.3580.957	DATE OF INSPECTION 04/19/2016
LOCATION OF INSTRUMENT (STREET AND CITY) 5434 S. Tower Drive, Battlefield, MO 65619		TIME OF INSPECTION 8:21 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 15120 EXP. DATE 04/29/2017

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34 SIMULATOR SN SD2251 SIMULATOR EXP DATE 08/12/2016

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102

TEST 2 .103

TEST 3 .101

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)		(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	--	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Ronald M Payne
TYPE II PERMIT NUMBER/EXPIRATION DATE 250192/08-18-2017	TELEPHONE NUMBER (417) 890-9876

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00242

Temp Date Time ^{s/} 210L

Air Blank:
04/19/16 20:21 .000
Calibration Check:
27 04/19/16 20:21 .102

Subject Name

ACC

Subject I.D.

Operator Name, I.D.

R. Payne / 1908

Location

Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00243

Temp Date Time ^{s/} 210L

Air Blank:
04/19/16 20:23 .000
Calibration Check:
27 04/19/16 20:23 .103

Subject Name

ACC

Subject I.D.

Operator Name, I.D.

R. Payne / 1908

Location

Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00244

Temp Date Time ^{s/} 210L

Air Blank:
04/19/16 20:25 .000
Calibration Check:
28 04/19/16 20:25 .101

Subject Name

ACC

Subject I.D.

Operator Name, I.D.

R. Payne / 1908

Location

Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00245

Temp Date Time ^{s/} 210L

Air Blank:
04/19/16 20:28 .000
Subject Test: Auto
28 04/19/16 20:28 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

R. Payne / 1908

Location

Battlefield P.D.

AS IV Serial no: 097460
Version no: 532B

TEST RECORD 00247

Temp Date Time ^{s/} 210L

VOID: RFI
12 04/19/16 20:34

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

R. Payne / 1908

Location

Battlefield P.D.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-1470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **15120** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **May 4, 2015**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1209%** (w/vol) ethyl alcohol. The expiration date for this lot number is **April 29, 2017** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:
Testing was conducted using Cerilliant Reference Standard lot number FN08051301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

RONALD M PAYNE

hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate perform field service and repairs,
 and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections
 7.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

ISSUE DATE 8/18/2015

MBER 250192

PIRES 8/18/2017

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

30-0771 (6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator PAYNE, RONALD
 Permit No 250192
 Date Issued 8/18/2015 Date Expires 8/18/2017